



WASHINGTON STATE PATROL

CODIS LABORATORY STR ANALYSIS PROCEDURES

CRIME LABORATORY DIVISION

February 2018

Table of Contents

Introduction	3
CODIS Case Approach	6
Quality Assurance/Quality Control	9
Performance Check and Calibration of Instruments	12
Processing of Convicted Offender Samples	14
DNA IQ™ Manual Extraction for Convicted Offender Samples	15
Microcon® Concentration of DNA	16
Quantifiler® Template SetUp	17
Quantifiler® – Reaction Preparation and 7500 Setup	18
Quantifiler® Data Analysis and Interpretation	19
BSD600 Duet Puncher Protocol	20
Fusion 6C Direct Amplification	22
Fusion 6C Extract Amplification	24
Amplification Product Preparation for the 3500xL	25
Running Plates on the 3500xL Genetic Analyzer	26
GENEMAPPER® ID-X SETUP	27
Guidelines for Evaluating DNA Profile Data – Fusion 6C	30
Guidelines for Evaluating Convicted Offender DNA Typing Profiles Using GeneMapper® ID-X V1.4 as an Expert System	37
CODIS Case File Content	39
Manual CODIS Case File Review With GeneMapper® ID-X V1.4	40
CODIS Case File TECHNICAL Review With GeneMapper® ID-X V1.4 as an Expert System	42
TempGenius™ Wireless Data Acquisition & Monitoring System	44
Reagent Preparation	45
TempGenius™ Wireless Data Acquisition & Monitoring System Maintenance	48
UV Irradiator Operating Instructions	49
BSD600 Duet Puncher Maintenance	50
ABI 7500 Real Time PCR System Maintenance	51
9700 Thermal Cycler Maintenance	52
3500xL Genetic Analyzer Maintenance	53
GeneMapper® ID-X V1.4 Server Maintenance	56
Tuttnauer Tabletop Autoclave – Instrument Maintenance	57
Appendix A: Administrative Procedures for Processing Offender Samples	59
Worksheets	64
References	65

INTRODUCTION

Procedure development is an ongoing process. Consult the DNA Technical Leader for significant changes to any of the following procedures or for developing additional procedures. Proposed changes shall be reviewed by those affected by this manual prior to incorporation. Allowances (non-significant changes) may be made for adapting procedures (e.g. volume, tube size/type, spin times, incubation times) to accommodate unusual samples and situations. Allowances to adapt a standard procedure are a recognized part of lab work, as they may be necessary to meet the requirements of certain samples. These deviations must be scientifically sound and shall be documented in the case file.

The standard method of analysis of DNA conducted by the Washington State Patrol CODIS Crime Laboratory (WSPCCL) for the convicted offender database is by the analysis of Short Tandem Repeat (STR) regions by Polymerase Chain Reaction (PCR). The WSPCCL examines 27 loci: the original Combined DNA Index System (CODIS) 13 core loci (CSF1PO, FGA, TH01, TPOX, vWA, D3S1358, D5S818, D7S820, D8S1179, D13S317, D16S539, D18S51, and D21S11); the expanded CODIS core loci (D1S1656, D2S1338, D2S441, D10S1248, D12S391, D19S433, and D22S1045); Amelogenin and DYS391 for sex discrimination; Penta D, Penta E, and SE33 for increased discrimination; and rapidly mutating Y STR loci DYS570 and DYS576.

PCR is a method used for the amplification of a specific DNA segment whereby two oligonucleotides (primers) anneal to opposite strands and flank a DNA region to be copied. The synthesis reaction is repeated for a number of cycles and results in the exponential accumulation of the specified DNA segment, the termini of which are defined by the 5' ends of the primers used in the reaction.

EXTRACTION

DNA suitable for PCR-based typing is currently obtained by the Promega DNA IQ™ extraction method, which uses paramagnetic resin to capture DNA.

HUMAN DNA QUANTITATION

The Applied Biosystems (AB) Quantifiler™ kits provide reagents and protocols necessary for the rapid and sensitive quantitation of human nuclear DNA. The role of the Quantifiler™ Human kit is to detect and quantify total human DNA. The genomic targets are PCR amplifiable sequence tagged sites on the human telomerase reverse transcriptase (hTERT) locus of chromosome 5. The procedure is quantitative PCR (qPCR) and is based on the real-time detection of a 5' fluorescent reporter dye that is cleaved and separated from a 3' non-fluorescent quencher incorporated on a DNA probe during PCR amplification. Also included in the kit is an internal PCR control (IPC) consisting of a non-naturally occurring DNA sequence that can detect the presence of PCR inhibitors. The qPCR is performed in the AB 7500 Real-Time PCR System instrument.

AMPLIFICATION

STR PCR amplification is performed using the AB GeneAmp® PCR System 9700 thermal cycler. The polymerase used requires heat activation. This has the advantage that the enzyme is not active during the setup of the samples. The formation of "primer dimers" is avoided and there is no time constraint placed upon the setup procedure. To activate the enzyme, the thermal cycler protocol includes a "hot start" at the beginning of the amplification process.

Three steps are involved in each amplification cycle. First, the template DNA is made single-stranded by heat denaturation. In the second step, the temperature is lowered so that annealing of the oligonucleotide primers to the template may occur. This interaction is favored because, initially, the ratio of primer to template is large. Annealing of the PCR primers to the template DNA is responsible for the technique's high level of specificity. Under optimized conditions, annealing of complimentary

DNA strands (primers) should only occur at those sites on the template flanking the region to be amplified. In this way, a sequence of several hundred bases can be selected from a background of, in the case of human DNA, 3×10^9 base pairs. In the third step, the temperature is raised to a point that favors both specific annealing of the primers and extension by the DNA polymerase. These three steps are repeated over and over again to achieve a sufficient quantity of amplified product. The target DNA is amplified at an exponential rate if one assumes a 100% efficient PCR reaction. Even when carefully controlled, a PCR reaction is rarely, if ever, 100% efficient.

The DNA polymerase adds an extra nucleotide (usually adenosine) to the end of a strand of DNA. If this reaction only occurs partially, then DNA of two lengths, differing by one base pair, will be produced from the same template and primers, causing the appearance of a split peak. Having amplified product of uniform size is important in STR analysis, as it is the length of the final product that is detected. A length variance of one base pair is detectable by the system, and therefore undesirable. To avoid this, the addition can either be prevented or conditions chosen so virtually all the amplification product is converted. Since the latter is easier, a 10-minute soak at 60°C is included at the end of the polymerization cycles to drive the addition reaction to completion.

Kits that contain all of the necessary components to amplify and detect Amelogenin and the 20 CODIS core loci are commercially available. It is possible to detect and distinguish the products of several loci amplified together. This is referred to as multi-component analysis. The PowerPlex® Fusion 6C kit allows for the co-amplification and detection of the 20 core loci, in addition to Amelogenin, Penta D, Penta E, SE33, DYS391, DYS570, and DYS576. This kit also allows for either the direct amplification of sample punches in PCR master mix, or for the amplification of DNA extracts with known concentrations.

Fusion 6C Kit Loci

Locus	Approx. Size Range (bp)	Dye Color
Amelogenin	80 – 89	Blue (FL-6C)
D3S1358	90 – 151	Blue (FL-6C)
D1S1656	152 – 209.5	Blue (FL-6C)
D2S441	211 – 252	Blue (FL-6C)
D10S1248	254 – 302.5	Blue (FL-6C)
D13S317	304.5 – 357	Blue (FL-6C)
Penta E	362 – 482	Blue (FL-6C)
D16S539	74 – 129.4	Green (JOE-6C)
D18S51	130 – 217.5	Green (JOE-6C)
D2S1338	221.5 – 304	Green (JOE-6C)
CSF1PO	313 – 366.5	Green (JOE-6C)
Penta D	373.5 – 470	Green (JOE-6C)
TH01	65 – 118	Yellow (TMR-6C)
vWA	121 – 192	Yellow (TMR-6C)
D21S11	197 – 266.5	Yellow (TMR-6C)
D7S820	268 – 315.5	Yellow (TMR-6C)
D5S818	317.5 – 380	Yellow (TMR-6C)
TPOX	390 – 448	Yellow (TMR-6C)
D8S1179	66 – 129.8	Red (CXR-6C)
D12S391	130.1 – 190.5	Red (CXR-6C)
D19S433	192 – 255	Red (CXR-6C)
SE33	270 – 429	Red (CXR-6C)
D22S1045	430 – 478	Red (CXR-6C)
DYS391	79.5 – 131	Purple (TOM-6C)
FGA	134 – 299	Purple (TOM-6C)
DYS576	302 – 370	Purple (TOM-6C)
DYS570	380 – 464	Purple (TOM-6C)
Internal Size Standard	60 – 500	Orange (WEN-6C)

DETECTION

Multicomponent Analysis

Multicomponent analysis uses two separate properties to distinguish between loci: length variance and fluorescent labels.

Length Variance

The amplified product from each locus will vary in length, dependent upon the number of repeats found in a particular STR. By careful selection of primers, it is possible to produce amplified products for several loci, all of which fall into discrete size ranges. Thus a particular locus can be identified by the size of its alleles.

Fluorescent Labels

Different fluorescent labels can be attached to the primers used in the amplification process. The emission spectra of these labels are sufficiently different such that the contribution made by each label in a mixture can be calculated both quantitatively and qualitatively. Thus, even if two separate DNA fragments co-migrate, they can be differentiated by their fluorescent labels.

By careful selection of both primers and fluorescent dyes, several different co-amplified loci can be distinguished based both upon their fluorescence and the size of the fragment lengths. Amplification kits are designed such that each dye employed detects loci in discrete size ranges, along with one dye reserved as an internal size standard.

Genetic Analyzers (CEs)

There are several instruments commercially available that will separate DNA fragments based upon their size and fluorescent labels. These instruments separate fragments using capillary electrophoresis. A window in the capillary is continually illuminated by a laser. As fragments of DNA with fluorescent labels pass by the window, they are excited and the resulting emitted light is collected and analyzed.

The WSPCCL currently uses AB 3500xL Genetic Analyzers.

Internal Size Standards

An appropriate internal size standard is added to each sample prior to electrophoresis. The internal size standard consists of several amplified fragments of DNA, all of which have been sequenced and are of known length. The migration time of each fragment is recorded and this data is used to calculate an estimated length for the other DNA fragments that are detected.

INTERPRETATION OF RESULTS

The data generated is analyzed using the software program, GeneMapper® ID-X, available from Applied Biosystems.

This program converts the raw data from AB Genetic Analyzers into discrete peaks, assigns a size to those peaks based upon their estimated length in base pairs, and converts the estimated fragment sizes in base pairs to designated alleles by comparing the fragment sizes to the known alleles in a standard called an allelic ladder. The program also displays that data in the form of an electropherogram.

CODIS CASE APPROACH

Quality in the CODIS Laboratory is maintained using the Crime Laboratory Division (CLD) Quality Operations Manual, the DNA Quality Assurance Manual, the FBI's DNA Quality Assurance Standards, and following the protocols in this manual.

ADMINISTRATIVE PROCESSING GUIDELINES

All convicted offender samples received in the laboratory are processed following the guidelines of the [Administrative Processing of Convicted Offender Samples](#) procedure. Convicted offender samples are collected by hundreds of collectors across Washington State using kits provided by the CODIS Laboratory. Kits are assembled in-house, and requests for these kits are handled by CODIS Laboratory staff. A detailed spreadsheet of kit requests is maintained by CODIS staff.

Completed kits are received through the mail or hand-delivery. Prior to acceptance by the lab, each completed kit is evaluated by a CODIS staff member. Information accompanying the DNA sample is assessed for completeness and accuracy. When a question arises about any of the submitted information, verification is done using the offender's criminal history record, court records, and/or contacting the collecting agency, prosecutor's office, or other law enforcement entity. Any information obtained using these sources is documented on the submission card along with the origin of the information and the date and initials of the individual recording the information. If a question arises about the quality of a sample, recollection can be requested. An electronic log is maintained of all rejected samples.

Each sample is assigned a unique laboratory number and is referenced only by that number thereafter. The electronic Convicted Offender Entry Form is used for data entry into the LIMS system. By using the SID of the convicted offender when entering data, the entry program is able to verify whether or not a DNA sample has already been received for that individual and will auto-populate the electronic entry form with the individual's name and date of birth. For any sample that is determined to be a duplicate, the request generated upon entry of the sample is automatically cancelled. Cancelled samples are retained, but are not typically typed unless needed for a hit confirmation or as a quality control check. For any sample without a SID, every attempt is made to determine if one exists. If one does not exist, the sample is entered manually using an acceptable alternative identification number (see [the Administrative Processing of Convicted Offender Samples](#) procedure for more details). When a non-duplicate sample is entered with a SID, the criminal history record of that individual is updated to indicate a sample has been received.

A barcode for each sample is generated once the sample has been entered into LIMS. Barcode labels are applied to any individual component of the sample submission.

Once processed administratively, offender samples are stored in a secure location in the CODIS Laboratory until ready to be typed.

LABORATORY PROCESSING GUIDELINES

All laboratory work is conducted following the procedures in this manual and the CLD DNA Quality Assurance Manual. Any deviation from laboratory procedure is documented in the case file and approved by the DNA Technical Leader or designee. The CODIS supervisor is designated by the DNA Technical Leader with the authority to approve acceptable minor deviations of STR interpretation guidelines and protocols. Before results may be released, peer review of a completed file is conducted in accordance with the CLD Quality Operations Manual, this manual, and the DNA Quality Assurance Manual by a qualified DNA CODIS analyst.

The date on top of the page following the Run Folder Review sheet is the date a sample set was started, usually the date samples were punched or prepared for extraction. The analyst will record the

date the case file was completed and reviewed by the analyst prior to technical review on the last page of the case file. Any changes made to the case file as the result of technical review will be documented with the analyst's initials and date to document that the change was made after the analysis complete date. For electronic records that do not automatically track changes, a note will be written in the case file describing the change, initialed, and dated.

As a quality control measure, sets may contain random reanalysis samples. These are known duplicate submissions for existing samples that were previously typed. The goal random reanalysis sample amount is five percent of samples per completed samples in a year. Random reanalysis samples may also include hit confirmation samples.

Convicted offender DNA samples are typically first processed in a 96-well plate format using direct amplification. Samples are punched into amplification plates with an automated puncher. The puncher includes a barcode reader, which scans each unique sample barcode and associates the sample with a specific well. When the barcode is unreadable, the sample number may be hand-entered. This sample to well association is maintained through analysis on a CE by copying the puncher-generated worklist data directly into an Excel workbook that automatically creates a CE template. The sample to well association is documented in the case file by saving the worklist tab as a PDF.

After amplification and running the samples and controls on a CE, the resulting data is analyzed using GeneMapper *ID-X*. This software is used to record information about samples that pass, but have quality flags or anomalies that the analyst must assess.

Samples that do not pass after the first round of analysis are re-tested using a method appropriate for the quality issue observed in the sample. Appropriate controls must also be run.

- PCR product can be diluted using TE buffer or injected for different times.
- As a quality control measure, manual manipulations from a 96-well plate to another plate or tube shall be witnessed when there is not sufficient genetic information in the first-pass data for comparison. The witness will document any steps observed by initialing and dating the applicable worksheet.
- Samples may be direct amplified again, but using a different combination of cycle number and amplification volume.
- Samples may also be extracted, quantitated, and amplified. Extracts can be cleaned up or combined and concentrated using Microcon filters in cases where samples are extremely weak or show signs of inhibition.

A manual punch can be used in lieu of the automated puncher when the batch of samples needing to be processed is reasonably small (e.g. samples that are being re-tested or a hit confirmation set).

When labeling tubes, the analyst will do so wither behind a shield, with a mask, or in a hood to prevent contaminating the tubes with their own DNA.

When a sample is consumed, this will be indicated on the outside of the sample card or envelope, along with the date and analyst's initials.

The samples are returned to the storage area once the analyst is done working with them.

Sample Tracking

CODIS analysts assign samples to themselves in LIMS prior to administrative review. In most cases, a LIMS worklist is made for each set of samples processed and a barcode identification of the LIMS

worklist is automatically generated. Following peer review, a mass-marking program allows the analyst to input the LIMS worklist barcode and enter the set number for Case Info, a result for Findings Entered, and the technical and administrative reviewers. Running the program updates each sample in the worklist with the input information.

For small sets of samples, tracking samples in LIMS may be done without a LIMS worklist. After assigning samples to themselves and analysis is complete, a result is added to Findings Entered, the set number is added to the Case Info tab, and the milestone is set to Draft Complete for each individual sample. The technical and administrative reviewers must also update the corresponding milestones when review is complete.

Once samples are marked "Admin Reviewed" in LIMS, any samples with SID numbers have the DNA Typed flag on the criminal history record updated.

Failed Samples

If a full profile is not generated for a sample after consuming the entire sample, the sample is considered to have failed. The request is cancelled in LIMS. A note about the sample failing or being SDIS-only is entered into the Notes field of the Case Info tab. If a duplicate is available that has not been typed, the analyst will un-cancel the request and type the duplicate sample at their next available opportunity. If the sample does not have a duplicate, the corresponding criminal history record is updated to reflect a "no" result for the DNA Taken and Typed fields. The collecting agency or other law enforcement entity is contacted to see if the individual is still incarcerated to recollect the sample. Whether or not a sample can be recollected and the source of the information is documented on the Run Folder Review worksheet. When an agency agrees to recollect a sample, the information is recorded on an electronic log for recollection requests. It is the responsibility of the individual requesting sample recollection to follow up with an agency to ensure recollection occurred.

QUALITY ASSURANCE/QUALITY CONTROL

DNA EXTRACTION AREAS

- All solutions used in DNA extraction are prepared according to the protocols in the Reagent Preparation section of this manual.
- All glassware to be used in preparation of critical reagents will be autoclaved. If a solution itself is to be autoclaved, the glassware used in its preparation need not be autoclaved beforehand. Pipette tips and microcentrifuge tubes that are sterilized (as in the manufacturing process) need not be autoclaved.
- Gloves will be worn where appropriate.
- Instruments used to cut or handle stains (scissors, forceps, scalpels, Harris punch) should be cleaned before each use.
- Use filtered pipette tips when pipetting any solutions.
- Care should be exercised when inserting pipettes into reagent containers. Stock reagents should be poured from the reagent bottle into a smaller container from which pipetting should be done. Avoid touching the container with the barrel of the pipette.
- The work area should be cleaned after each use with 10% bleach or other appropriate cleaner.
- The interiors of microcentrifuges should be cleaned periodically.
- All non-sterile plates and tubes will be autoclaved and UV irradiated prior to use.

PCR SET-UP AREA

- Gloves will be worn when working in the PCR set-up hoods.
- Set up PCR according to protocol. Use only the pipettes dedicated to the PCR set-up area. Use sterile, filtered pipette tips. Dispose of pipette tips in a waste container in the hood.
- The hood will be cleaned on a regular basis with a 10% bleach solution (or other appropriate cleaner) and then rinsed with water (if using bleach). Pipette barrels will also be wiped regularly with a 10% bleach solution (or other appropriate cleaner). Additionally, the UV light in the hood (if so equipped) will be utilized after each use of the hood.
- All non-sterile plates and tubes will be autoclaved and UV irradiated prior to use.

POST-AMPLIFICATION AREAS

- All solutions used in the post-amplification areas are prepared according to the reagent manual protocol.
- Gloves will be worn where appropriate.
- Only dedicated lab coats will be worn when working in this area.
- Handling of PCR products will be done with dedicated pipettes and filtered pipette tips.
- If bench top protective paper is used, it will be changed at least weekly.
- Any PCR product or other items contaminated with PCR product will be discarded in the appropriate container in the post-amplification areas.
- PCR product may only be stored in a dedicated refrigerator or freezer.

CONTROLS AND STANDARDS

Reagent Blank

A reagent blank will be prepared each time a set of DNA samples is extracted. The reagent blank will be prepared like the other samples being extracted, except no source of DNA will be added to this sample. Any generated reagent blank shall be amplified using the same primers, instrument model, and concentration conditions as required by the sample(s) containing the least amount of DNA. The amplified reagent blank shall be typed using the same instrument model, injection conditions, and most sensitive volume conditions of the extraction set.

In accordance with FBI QAS Standard 9.5.3.2, if multiple reagent blanks are used within an extraction set, each reagent blank shall be quantified and at least one shall be amplified if any of the specimens associated with the extraction set will be amplified. If multiple reagent blanks are used and quantified with an extraction set, at minimum, the reagent blank that demonstrates the greatest signal, if any, shall be amplified and characterized.

When concentrating samples where a single extraction generated multiple reagent blanks, only one of those reagent blanks is required. The reagent blank taken forward should be one that, if formerly quantified, demonstrated the greatest signal, if any, unless that reagent blank was already demonstrated to be contaminated.

If the reagent blank associated with the extraction set or sample being amplified is depleted, an analyst shall not continue. The volume of sample amplified cannot exceed that of its reagent blank.

Internal Laboratory Control

The internal laboratory control is a sample from a known individual within the laboratory system who does not work in the CODIS Laboratory. This control is not required, but, when used, it should be treated as a regular sample. Every attempt should be made to get the control to successfully genotype, but if it is the only sample delaying the completion of a set, testing may cease with an explanation provided in the case file.

Should a positive amplification control fail, an internal laboratory control that is run in the same amplification may be used in its place to demonstrate the amplification worked as expected.

Positive and Negative Amplification Controls

At the amplification step, two extra tubes containing the amplification master mix will be prepared. To one tube the appropriate volume of positive control DNA will be added, and to the other an equal volume of TE or Amplification Grade Water will be added.

A positive and negative amplification control will be included in every set of amplifications conducted and the negative control will be processed under the same conditions as the positive control.

For direct amplification reactions, the negative amplification control is considered a combined reagent blank and negative amplification control. It shall be typed using the same instrument model, injection conditions, and most sensitive volume conditions used for the samples in the batch.

Internal Size Standards

An appropriate internal size standard will be included with every sample prepared to be run on the CE. The data collected from this standard will be used to verify the quality of a particular injection and also be used to estimate the size of any DNA fragments present in the run. It is essential to have the data from this standard in order to process samples in the GeneMapper® ID-X program.

Allelic Ladder

An appropriate ladder will be included with every set of samples run on the CE. A passing allelic ladder is essential in order to process samples in the GeneMapper® ID-X program.

QUALITY CONTROL OF REAGENTS

The following critical reagents shall be evaluated prior to use:

- Promega DNA IQ™ kit
- Dithiothreitol
- TE

- Quantifiler™ kit
- Promega G147A for the QF standard curve
- PowerPlex® Fusion 6C kit
- PunchSolution™ kit

REAGENT QUALITY CONTROL TESTING

Unless a quality control procedure indicates otherwise, a known DNA sample is used to perform the QC of a new lot number of reagent or kit before it may be used on convicted offender samples. The known DNA sample must be analyzed using the current laboratory protocols. The known DNA sample must type correctly and meet all data analysis parameters as required in the current laboratory protocols. If the sample does not type correctly or does not meet the analysis parameters, the reagent or kit may not be used for convicted offender analysis until the problem is resolved. Scientifically valid exceptions pertaining to meeting the analysis parameters can be made if approved by the CODIS supervisor in consultation with the DNA Technical Leader. The successful completion of the QC is documented on the reagent or kit and on the appropriate QC worksheet. All QC testing will be verified by another CODIS scientist to ensure that analysis parameters are met, the paperwork is filled out correctly, and the electronic data is present. This verification will be documented on the QC worksheet, which will then be saved as a PDF and stored with the electronic data.

PERFORMANCE CHECK AND CALIBRATION OF INSTRUMENTS

Routine maintenance procedures are described by instrument/equipment elsewhere in this manual (see Table of Contents). A maintenance item does not necessarily have to be performed on-schedule for an instrument that isn't being used at the time the maintenance is due. The operator will ensure all necessary maintenance items are performed prior to use. If annual maintenance is required of an instrument, it is not necessary to perform it within the year if the instrument is taken off-line.

The performance check and/or calibration of instruments demonstrates they are working properly and will produce the expected results. Performance checks will be conducted at least annually, and for critical instruments after calibration, maintenance, or repair. Written documentation of calibrations and performance checks will be maintained in the CODIS laboratory.

- pH meters will be calibrated with known pH standards according to the manufacturers' instructions before they are used.
- Balances will be checked at least annually by an external provider.
- For each refrigerator/freezer/oven/water bath/heat block, variance should be established as per the [Equipment Calibration and Maintenance](#) section of the DNA Analysis Quality Assurance manual. If, when the temperature is checked, the equipment is functioning outside that range, use of the equipment in forensic testing should be stopped immediately and measures taken to return the equipment to function within its accepted range.
- The operating temperature of refrigerators, freezers, and any other storage unit or area used to store reagents or samples that would degrade if not maintained at or below a given temperature will be recorded each working day or monitored using the TempGenius system.
- The operating temperature of ovens, water baths, or other apparatus that is used to maintain a constant temperature of reagents or a reaction should be checked each working day, prior to its use, or monitored using the TempGenius system.
- Semiannually, the thermometers/probes used to record these temperatures will be verified against a NIST traceable thermometer. The thermometer/probe can remain in service if it is within $\pm 4^{\circ}\text{C}$ of the NIST traceable thermometer. Alternatively, a thermometer can be used that is NIST traceable and certified for a specific time period as long as it is replaced or re-certified before that period has expired.
- The operating temperatures of heat blocks for direct amplification pre-treatment and for denaturing/snap-cooling amplification product do not require daily recording of temperatures. The thermometers of these units will be periodically verified against a NIST traceable thermometer as described above.
- NIST traceable thermometers will be re-certified annually by an appropriately qualified external agency.
- Thermal cyclers are a critical instrument and will be performance checked semi-annually using the appropriate temperature probe and the directions provided by the manufacturer of the thermal cycler.
- Annually, the probes used to check the calibration of thermal cyclers will be re-certified by an appropriately qualified external agency.

- Pipettes will have their calibration checked annually by a qualified external agency.
- The following instruments are critical instruments and will be performance checked prior to use in convicted offender sample analysis and following maintenance or repair by qualified technicians:
 - BSD600 Duet Puncher
 - AB 7500 Real-Time PCR System
 - AB GeneAmp PCR System 9700 thermal cycler
 - AB 3500xL Genetic Analyzer

PROCESSING OF CONVICTED OFFENDER SAMPLES

DNA samples received by the CODIS Laboratory are treated as reference materials.

Buccal samples on FTA paper (from a WSP provided kit) from convicted offenders are collected and submitted to the Crime Laboratory by law enforcement facilities. In rare instances, blood samples are drawn instead. It has been requested that these samples be submitted in EDTA vacutainer tubes, accompanied by a Washington State Patrol Crime Laboratory Division Convicted Offender DNA Data Sheet (WSP-CL-549).

On receipt, each whole blood sample is prepared as follows:

1. Ten to twelve 75 µL stains are prepared on Schleicher & Schuell 903 paper or other clean filter paper. The stains are air-dried, sealed in plastic bags, and frozen at -20°C to -80°C, pending analysis.
2. Any remaining whole blood is destroyed by pouring into a disinfectant solution. The destruction of each blood sample is witnessed.

DNA typing of the convicted offender blood samples is done using a portion of one of the prepared 75 µL bloodstains. Similarly, DNA typing of convicted offender buccal samples is done using a portion of the submitted FTA card.

DNA IQ™ MANUAL EXTRACTION FOR CONVICTED OFFENDER SAMPLES

This method is for DNA extraction from buccal samples on FTA® paper and blood stains on filter paper typically provided as samples from convicted offenders. This method may also be used for swabs, foam-applicators, and gauze or other cloth materials. The DNA IQ™ extraction system (Promega) is a DNA extraction method that uses a paramagnetic resin to capture DNA. Provided there is sufficient DNA present to saturate the resin, elution into 100 µL will give a concentration of approximately 1 ng/µL. A range of elution volumes can be used to vary the concentration of the resulting DNA extract. Depending on the nature of the sample being extracted, the elution volume or the amount of sample extracted can be adjusted.

1. Place sample into a 2 mL tube basket assembly or, for smaller sample sizes, a 1.7mL microcentrifuge tube (or other tube/tube basket assembly that has been performance tested and found to be suitable).
2. Add 150 µL (or up to 300 µL for larger sample amounts) prepared Lysis Buffer +DTT and incubate at 95°C for 30 minutes.
3. Using a sterile technique, remove any punches from the tubes. If using a tube/basket assembly, transfer sample into the basket insert and place the insert back into the tube. Spin at maximum speed in a microcentrifuge for 2 minutes. Remove and discard the basket insert, including the basket contents.
4. Vortex stock resin bottle for 10 seconds. Add 3-7 µL re-suspended resin to sample. Keep stock resin re-suspended while dispensing to obtain uniform results. Vortex sample tube for 3 seconds and incubate at room temperature for 5 minutes.
5. Vortex sample tube for 2 seconds. Place sample tube on magnetic stand and carefully discard solution with a disposable pipette without disturbing resin.
6. Add 100 µL prepared Lysis Buffer +DTT and vortex tube for 2 seconds. Place tube in magnetic stand. Carefully remove and discard solution.
7. Add 100 µL prepared 1X Wash Buffer to tube and vortex for 2 seconds. Place tube in magnetic stand and carefully discard solution. Repeat for a total of three washes.
8. With lids open, air-dry 5-20 minutes at room temperature on the magnetic stand.
9. Add Elution Buffer (between 25 and 100 µL; usually 50 µL) and close lids. Vortex tube for 2 seconds and incubate at 65°C for 5 minutes.
10. Remove tubes from heat, immediately vortex for 2 seconds and place on magnetic stand. Remove DNA extract to a microcentrifuge tube for immediate use and/or storage. Samples may be stored in a refrigerator for a few days or in a freezer for long term storage.
- 11.

MICROCON® CONCENTRATION OF DNA

This procedure can be used to combine and concentrate weak extracts or to clean up extracts that show signs of inhibition (e.g. poor amplification of the IPC during quantitation, ski slope effect in the electropherogram). When using a Microcon, the associated reagent blanks shall be processed concurrently with the sample(s). The Microconned reagent blank will be named using the date of the procedure.

1. Assemble a Microcon® Fast Flow concentrator unit. To the top of the concentrator, add 30-100 µL TE, if desired. Transfer the DNA extracts to be concentrated to the top of the concentrator. The total volume of extract and TE added to the sample reservoir should be no more than 500 µL.
2. Cap the concentrator and spin in a centrifuge at ~500 to 5000 x g for about 10-15 minutes or longer (settings can be adjusted according to centrifuge type and sample viscosity).
3. Carefully remove the concentrator unit from the assembly and discard the filtrate fluid from the filtrate vial. Return the concentrator to the top of the filtrate vial.
4. Add 200 µL TE to the concentrator. Recap and spin the assembly in the centrifuge for ~10 minutes at ~500 to 5000 x g.

NOTE: Additional washes may be required to remove inhibitors that may be present and should be done on samples of extremely limited quantity.
If the microcon is being used to just concentrate or reduce sample volume, the addition of TE as a wash step is not necessary.

5. Remove the cap, if adding TE; add the desired final volume of TE to the concentrator.
6. Remove the concentrator from the filtrate vial and carefully invert the concentrator onto a new labeled retentate vial. Discard the filtrate vial.
7. Centrifuge the assembly at 1000 x g for ~5 minutes (or pulse briefly).
8. Discard the concentrator. Cap the retentate vial.

Samples are ready for human DNA quantitation. Use samples immediately for PCR or refrigerate samples for storage while the sample is actively being worked (avoid repeated freeze-thaw cycles). Freeze samples for long term storage. Prior to use for PCR, vortex and pulse spin samples.

QUANTIFILER® TEMPLATE SETUP

Prior to quantitation, a Quantifiler® template is set up and used to assign an imported plate document with the appropriate cycling conditions, dyes, reaction volume, and analysis settings.

1. Open the 7500 Software and select Advanced Setup from the main window.
2. Select Experiment Properties. Add the name of the 7500 Software version in the Comments field. Ensure that the following properties are selected:
 - a. 7500 (96 Wells)
 - b. Quantitation – Standard Curve
 - c. TaqMan® Reagents
 - d. Standard (~2 hours to complete a run)
3. Select Plate Setup:
 - a. Rename “Target 1” as “Sample”. Ensure the Reporter dye is “FAM” and the Quencher is set to “None”.
 - b. Add a new target and rename it “IPC”. Ensure the Reporter dye is “VIC” and the Quencher is set to “None”.
4. Select Run Method and set the following parameters:
 - a. Reaction Volume Per Well: 10 µL
 - b. Holding Stage
 - Step 1: 95.0°C for 10:00
 - c. Cycling Stage (Number of Cycles = 40, do not Enable AutoDelta)
 - Step 1: 95.0°C for 00:15
 - Step 2: 60.0°C for 01:00
5. Select Analysis → Analysis Settings. Select each target (“Sample” and “IPC”) and set the following parameters:
 - a. Uncheck Default C_T Settings
 - b. Uncheck Automatic Threshold. Set the Threshold to 0.2.
 - c. Uncheck Automatic Baseline. Set the Baseline Start Cycle to 6, and the End Cycle to 15. Select Apply Analysis Settings
6. Select File → Save As Template. Save as “QF template”.

QUANTIFILER® – REACTION PREPARATION AND 7500 SETUP

1. Calculate the volumes of Quantifiler® PCR Reaction Mix (5.3µL per sample/standard) and Primer Mix (4.5 µL per sample/standard) needed to prepare the reaction master mix.
2. Thaw the Primer Mix completely, vortex 3 to 5 seconds and pulse spin briefly before opening the tube.
3. Swirl the PCR Reaction Mix gently before using. Do not vortex.
4. Pipette the calculated volumes of components into an appropriately sized tube.
5. Vortex the master mix for 3 to 5 seconds, then pulse spin briefly.
6. Obtain a 96-well reaction plate and seat in the optical support base. The reaction plate shall not be rested on any surface without the base. The plate shall not be touched on the bottom of the wells.
7. Dispense 9.2 µL of the master mix into each reaction well of the reaction plate.
8. Add 0.8 µL of sample, standard, or control to the appropriate well of the reaction plate, as determined by the plate setup sheet.
9. Seal the reaction plate with an optical adhesive cover. Care shall be taken not to touch the cover. The plastic applicator shall be used to seal the cover to the reaction plate. Drag the applicator across the cover several times to ensure a proper seal. Seal around the edges and remove the white strips from the cover.
10. Centrifuge the plate as necessary to eliminate bubbles.
11. Open the 7500 and place the 96 well plate in the precision holder with well A1 in the upper left corner. Push the tray closed.
12. Open the 7500 Software and select Template from the main window. Select “QF template”.
13. Select File → Import. Browse for and select the desired plate document (txt).
14. Select Start Import. When prompted for continuing with the import, select Yes.
15. Verify that the plate layout appears as it should, and that all quantitation parameters are appropriate.
16. Select File → Save As. Save the file (eds) in a designated results folder
17. Select Start Run.

QUANTIFILER® DATA ANALYSIS AND INTERPRETATION

The ABI Quantifiler® Human DNA Quantification Kit and 7500 Real-Time PCR System are used to determine the amount of PCR amplifiable human DNA. The results can aid in determining whether inhibition is present in a sample, if sufficient DNA is available for STR analysis, and how much DNA to target for amplification.

1. Open the completed .eds quantitation file and verify the Run Method and Analysis Settings are correct (see Quantifiler® Template Setup).
2. Under the Analysis menu, select Standard Curve. Review the following values:
 - a. **Slope:** indicates amplification efficiency
 Typical Validated Range -2.9 to -3.3
 Average: -3.1
 Ideal (100% efficiency): -3.32

 If the slope is outside the determined range then a check of the assay setup, software setup, reagents, and instrument may be appropriate. Under Experiment Properties, indicate in the Comments field that the slope is outside of the validated range and that results should be used with caution.
 - b. **R² (correlation coefficient):** indicates the statistical significance of the standard curve. An R² value greater than 0.98 is desirable. If R² is less than 0.98, consider the preparation and efficacy of the quantitation standards. Under Experiment Properties, indicate in the Comments field that R² is less than 0.98 and that results should be used with caution.
3. Under the Analysis menu, select Amplification Plot to review the Internal PCR Controls (IPCs). The IPCs in the assay monitor for the presence of PCR inhibitors and may indicate when sample reprocessing is necessary. The normal C_T (threshold cycle) for an IPC should range from 20 to 30.

Sample/Standard (FAM dye)	IPC (VIC dye)	Interpretation
No amplification	Amplification	True negative
No amplification	No amplification	Cannot distinguish between inhibition and absence of DNA
High concentration with low C _T	No amplification	IPC efficacy may be suppressed by high DNA concentration
Weak amplification with high C _T	No amplification	Partial PCR inhibition

4. In the Amplification Plot section, click on View Well Table tab to review the sample quantitation results. The Quantity column displays concentration in ng/μL.

BSD600 DUET PUNCHER PROTOCOL

The BSD600 Duet Puncher punches 1.2 mm discs from dried sample cards (buccal and blood samples on FTA or filter paper) and deposits them into 96-well plates for direct amplification. The software creates an electronic worklist that is used for sample tracking.

CREATING A TEST

Tests serve as templates for standard punching protocols. Multiple tests can be created in advance for different needs, such as full plates of convicted offender samples or proficiency tests.

1. Launch the BSD600 Duet Software and log in to display the main menu.
2. Click on Configure System. Ensure that:
 - a. File tab: the desired worklist format is selected.
 - b. Punching tab: number of cleaning strikes and number of samples between cleaning are set to 1.
3. Click on the Edit Test Sequences button, then Create a new test. Select the appropriate Plate definition in the dialog box (Microtitre.try for a 96-well plate).
4. Select Test > Test configuration. Click on each tab:
 - a. Position settings may be set based on user preference.
 - b. Automatic filling should be enabled, beginning at the top left corner and filling vertically. Wells per sample should be set to 1.
 - c. Spot per Well should be set to 1 for each small punch sample type and 0 for each large punch sample type.
 - d. Plate Barcodes should be left empty or set to allow any type of barcode.
 - e. End Controls should be set to 0.
5. Wells are set to contain Samples by default. To change the type of well, double click and select the desired sample type from the menu. Select Leave Empty for any wells that should not contain a punch.
6. Select Test > Standard & control names. If desired, different standards and controls may be given more specific names.
7. An additional plate may be added to the test from the Plates menu if needed.
8. Save the test using a unique name.

OPERATING THE PUNCHER

1. Turn on the humidifier, ideally 30 minutes prior to punching.
2. If samples are to be punched into a reagent for direct amplification (such as PunchSolution or amplification master mix), add reagent to the plate(s).
3. Turn on the instrument and vacuum system.
4. Launch the BSD600 Duet Software and log in.

5. Select the Distribute Spots Program option and choose an appropriate test.
6. Load plate(s) after the puncher moves into position, as directed by the software.
7. If desired, turn on the Auto-trigger and set to a time interval comfortable for the user. Alternatively, the foot switch may be used to activate the punching process.
8. Scan the barcode of the first sample to be punched, which will be displayed on the computer screen. If a sample does not have a barcode, the sample number can be typed into the software. Controls are pre-named based upon the test chosen.
9. Slide the card beneath the card clamps so that the red light targeting pattern can be seen on the card. This indicates the precise position from which a disc will be punched.
10. Move the card to the desired punching location. If the Auto-trigger is turned on, the punching process will occur automatically based on the user-set time interval. Otherwise, push the foot switch to activate punching.
11. After the punch has been deposited in the plate, the software will prompt the user to punch a filter paper waste punch following the same process.
12. The punching cycle is repeated until all samples for a test are punched.
13. Remove plates from the puncher.

FUSION 6C DIRECT AMPLIFICATION

Direct amplification can be used for buccal and blood samples on FTA or filter paper. 1.2 mm punches are pre-treated with PunchSolution to maximize the amount of DNA available for amplification. An appropriate PCR cycle number should be chosen based on the sample type being amplified. Because only one punch is used for direct amplification, the reactions are highly dependent on sampling, especially for non-uniform buccal on FTA samples. For re-testing, cycle number and reaction volume can be varied to obtain a useable profile. In general, increasing cycle number results in higher peaks, while a larger reaction volume tends to produce a more balanced profile. For very weak samples, the analyst should consider alternative case approaches, such as concentrating DNA extracts for a standard amplification.

1. Dispense 10 µL PunchSolution into each tube/well, including the negative amplification control. Exclude any positive amplification control tubes/wells.
2. Place one 1.2 mm punch per sample into each tube/well and incubate at 70°C for 30 minutes or until tubes/wells are dry. The plate must be left uncovered for evaporation to occur.
3. Prepare the components of the Fusion 6C amplification kit.
 - a. At first use, thaw the 5X Master Mix, 5X Primer Pair Mix and Amplification Grade Water completely. After first use, store at 2 – 10°C.
 - b. Centrifuge the 5X Master Mix and 5X Primer Pair Mix tubes briefly to bring the contents to the bottom, then vortex the reagents for 15 seconds before each use. Do not centrifuge after vortexing, as this may cause a concentration gradient.
 - c. Vortex and briefly spin the 5X AmpSolution Reagent. If the reagent is turbid, it may be warmed slightly at 37°C, then vortexed until clear.
 - d. Appropriately dilute the 2800M positive control (+C) DNA with Amplification Grade Water so that 1 µL of +C solution contains the desired DNA target. 2800M is provided at 10 ng/µL. Suggested targets:

Sample Type	Target
Buccal	2 ng
Blood	5 ng

4. Determine the number of reactions to set up (# samples = n), including positive and negative controls. A 5% buffer is included in the amount of each reagent to compensate for volume loss during pipetting.

Component	Half Volume	Full Volume
Amplification Grade Water	5.25 µL x n	10.5 µL x n
5X Master Mix	2.63 µL x n	5.25 µL x n
5X Primer Pair Mix	2.63 µL x n	5.25 µL x n
5X AmpSolution Reagent	2.63 µL x n	5.25 µL x n

5. Set up the amplification. The following steps should be performed in a biological hood or other area designated for PCR set up.

- a. Prepare the master mix using the table in step 4. Vortex for 5 – 10 seconds and dispense into each tube/well – 12.5 µL for the half volume reaction or 25 µL for the full volume reaction.
 - b. Vortex and spin the diluted 2800M DNA, then add 1 µL to any positive control tubes/wells.
6. Seal or cap the tubes/plate and, if needed, briefly centrifuge to bring the punches and master mix to the bottom.
7. Load the tubes/plate onto a thermal cycler and start the appropriate amplification protocol. Ensure that Max Mode is selected for ramp speed and the correct reaction volume is input.

THERMAL CYCLING PROTOCOL

96°C for 1 minute, then:

96°C for 5 seconds
60°C for 1 minute
for desired # of cycles, then:

60°C for 10 minutes

4°C for ∞

Reaction Volume	Sample Type	Cycle #
12.5 µL	Buccal	24 – 27
	Blood	23 – 25
25 µL	Buccal	24 – 26
	Blood	23 – 25

FUSION 6C EXTRACT AMPLIFICATION

The target amplification range of template DNA is approximately 0.125 ng to 4.0 ng based on quantification results. For samples below this DNA target amount, the analyst should consider alternative case approaches, such as combining and concentrating DNA extracts.

1. Prepare the components of the Fusion 6C amplification kit.
 - a. At first use, thaw the 5X Master Mix, 5X Primer Pair Mix and Amplification Grade Water completely. After first use, store at 2 – 10°C.
 - b. Centrifuge the 5X Master Mix and 5X Primer Pair Mix tubes briefly to bring the contents to the bottom, then vortex the reagents for 15 seconds before each use. Do not centrifuge after vortexing, as this may cause a concentration gradient.
 - c. Appropriately dilute the 2800M positive control (+C) DNA with Amplification Grade Water or TE buffer to the desired template amount. 2800M is provided at 10 ng/μL.
 - d. A negative amplification control (-C) will be run with every amplification. The -C will be prepared using either Amplification Grade Water or TE buffer depending on which reagent was used to normalize the samples being amplified.
2. Determine the number of reactions to set up (# samples = n), including controls. A 5% buffer is included in the amount of each reagent to compensate for volume loss during pipetting.

Component	Volume
5X Master Mix	5.25 μL x n
5X Primer Pair Mix	5.25 μL x n

3. Set up the amplification. The following steps should be performed in a biological hood or other area designated for PCR set up.
 - a. Prepare the master mix using the volumes above. Vortex for 5 – 10 seconds and dispense 10 μL into each tube.
 - b. Add 15 μL of DNA extract and 2800M DNA diluted to the desired target with either amplification grade water or TE buffer to the appropriate tubes. Add Amplification Grade Water or TE buffer to the negative amplification control tube.
4. Load the tubes onto a thermal cycler and start the appropriate amplification protocol. Ensure that Max Mode is selected for ramp speed and the 25 μL reaction volume is input.

THERMAL CYCLING PROTOCOL

96°C for 1 minute, then:

96°C for 5 seconds

60°C for 1 minute

for 29 cycles, then:

60°C for 10 minutes

4°C for ∞

AMPLIFICATION PRODUCT PREPARATION FOR THE 3500XL

Formamide is a teratogen – wear gloves.

1. Prepare a mixture of Hi-Di™ Formamide and WEN ILS 500.

- $[(9.0 \mu\text{L formamide}) \times \# \text{ samples}] + [(1.0 \mu\text{L WEN}) \times \# \text{ samples}]$
Example: 1000 μL formamide + 111.1 μL WEN

The loading cocktail may also be prepared using a concentration down to 9.75 μL formamide + 0.25 μL WEN. This will decrease the intensity of the size standard peaks.

2. Vortex for 10 – 15 seconds and spin briefly. If making a stock tube, label with the WEN lot number, date of preparation, and analyst's initials.
3. Dispense 10 μL of formamide/WEN mixture to the appropriate wells of the CE plate.
4. Add 1 μL of PCR product or allelic ladder to each well. PCR product may be diluted with TE buffer up to 1:20 as needed.
5. Cover the plate with a septum and spin briefly to remove bubbles.
6. Denature samples at 95°C for 3 minutes, then snap-cool for 3 minutes.
7. Place the plate in a plate base and cover with a plate retainer.

RUNNING PLATES ON THE 3500XL GENETIC ANALYZER

1. On the Dashboard of the 3500 Series Software, click the 'Start Pre-Heat' button to turn on the oven.
2. Complete any maintenance that is required.
3. With the instrument door closed, push the tray button to bring the autosampler forward. Place any plates to be run on the autosampler. When the door is shut, the autosampler should return to its home position.
4. Navigate to the Main Workflow portion of the 3500 Series Software to create a new plate or import a plate from a template. The following parameters must be applied to each plate:
 - a. Assay: 6sec_F6C, 12sec_F6C, 18sec_F6C, 24sec_F6C
 - b. Results Group: CODIS
 - c. File Name Convention: CODIS
5. On the 'Load Plate for Run' screen, verify that any plates on the instrument are linked correctly and start the run.
6. Wait until the 'Monitor Run' screen is displayed before leaving the instrument. Error messages (ex. expired reagent installed, not enough POP-4 remains for all injections, etc.) may appear before this screen is displayed; the run will not proceed until the messages are acknowledged.

GENEMAPPER® *ID-X* SETUP

The following procedure applies primarily to *ID-X* users with Admin profile access and shall be applied to any Server copy of *ID-X*. All users will set up project options and CODIS table export settings on their *ID-X* Clients.

If an individual user name has not been created, log in as user “gmidx”. Otherwise, log in using the individual user name. Once an Admin profile user has been created, any changes to the software shall be done from that profile to maintain a more accurate audit trail record.

SETTING UP A USER ACCOUNT

1. Open Security Manager
2. Select the appropriate user account type:
 - a. For an administrator, select Databasing Tech Lead
 - b. For an analyst, select Databasing Analyst
3. Select Edit > Duplicate
4. In the name field, change the name to the CODIS login of the individual the account applies to.
5. Enter the full name of the individual under Full Name in the User Details section.
6. Select Pre-Expire for the password to prompt the new user to create a password upon their first login.
7. Select Set Password and create a generic password for the new user.
8. Select File > Save

For custom user groups, user accounts, or security groups, follow the steps as outlined in the GeneMapper *ID-X* Administrator's Guide. Additional information about editing users can also be found in the Administrator's Guide. Only an Admin level user with the input of the CODIS supervisor and/or the DNA Technical Leader may make these changes.

After any changes are made in the Security Manager, a copy of the settings should be exported as a backup. If more than one Server exists, the settings shall be imported to that computer to ensure they are the same (see the GeneMapper *ID-X* Administrator's Guide for instructions on exporting and importing security settings).

SETTING UP THE AUDIT TRAIL CONFIGURATION

1. Open the *ID-X* software
2. Choose Admin > Audit Manager > Setting
3. Log in to Audit Manager
4. Choose “Allele” from the left navigation pane
 - a. Set “deleted” and “created” to “on”
 - b. Set “modified” to “silent”
5. Choose “Sample” from the left navigation pane
 - a. Set “deleted”, “OverrideSQ”, “modified, and “created” to “off”
 - b. Set “OverrideGQ” and “OverrideCGQ” to “silent”
 - c. Set “ALLELE” to “on”

SETTING UP THE ANALYSIS METHOD FOR CONVICTED OFFENDER ANALYSIS

1. Open GeneMapper *ID-X* Manager.
2. Select the Analysis Method tab and click New.
3. Select the General tab and enter the following:
 - a. Name: F6C_3500xL_<implementation date>
 - b. Security Group: Databasing Security Group
 - c. Instrument: 3500xL

4. Select the Allele tab and enter the following:
 - a. Bin Set: PowerPlex_Fusion_6C_Bins_IDX_v1.2
 - b. Use marker-specific stutter ratio and distance if available: check box
 - c. Ensure that all other values are set to 0.0
5. Select the Peak Detector tab and enter the following:
 - a. Ranges:
 - i. Analysis: Full Range
 - ii. Sizing: Partial Sizes, 60 – 500
 - b. Smoothing and Baseline: Light Smoothing, 51 pts Baseline Window
 - c. Size Calling Method: Local Southern Method
 - d. Peak Amplitude Threshold: 120 for all dye colors
 - e. Min. Peak Half Width: 2 pts
 - f. Polynomial Degree: 3
 - g. Peak Window Size: 15 pts
 - h. Slope Threshold: start and end at 0.0
 - i. Normalization: uncheck the box
6. Select the Peak Quality tab and enter the following:
 - a. Homozygous min peak height: 500.0
 - b. Heterozygous min peak height: 120.0
 - c. Max Peak Height (MPH): 50000.0
 - d. Min peak height ratio: 0.60
 - e. Max peak width: 1.5
 - f. Max expected alleles: 2 for autosomal & AMEL, 1 for Y markers
 - g. Allelic Ladder Spike: Enable detection, Cut-off Value 0.2
 - h. Sample Spike Detection: Enable
7. Select the SQ & GQ Settings tab and enter the following:
 - a. AMEL Cross Check (ACC): 0.5
 - b. Do not alter the remaining default settings
8. Save the analysis method.

SETTING UP PROJECT OPTIONS

1. From the Project Window, go to File > Project Options.
2. Select the General tab and choose the following:
 - a. Under Project, check Open Blank Project
 - b. Under Data Access Control, select Databasing Security Group
3. Select the Add Samples tab and choose the following:
 - a. Set Analysis Method to: Current Analysis Method
 - b. Set Size Standard to: WEN_ILS_500_CS
 - c. Set Panel to: PowerPlex_Fusion_6C_Panels_IDX_v1.2
4. Select the Analysis tab and choose the following:
 - a. Under Analysis Summary check:
 - i. Stop analysis and display Analysis Requirements Summary
 - ii. Stop analysis and display Allelic Ladder Analysis Summary
 - iii. View Analysis Summary
 - b. Under Quality Metrics Display, check Symbols
5. Click OK to save the Project Options.

IMPORTING LAB REFERENCE AND CUSTOM CONTROL PROFILES

1. From the Project Window, open the run folder containing the desired profile
2. Go to the Samples tab and ensure the Sample Name is the desired custom name for the reference or control
3. Select the sample and go to Tools > Add Profile > Lab Reference (in the case of a reference) or Custom Control (in the case of a control)
4. Click Close in the Add Profile Results dialogue box to save the profile

CODIS TABLE EXPORT SETTINGS

- 1. Go to Tools > CODIS Export Manager
- 2. Under Source Lab IDs, enter the source lab and select add
- 3. Under Destination Lab IDs, enter the destination lab and select add
- 4. Click OK
- 5. Navigate to the CODISMarkers file and ensure that all loci in the amplification kit being user are included. Refer to GeneMapper *ID-X* Software Help for details on location and editing of the file.

GUIDELINES FOR EVALUATING DNA PROFILE DATA – FUSION 6C

The following guidelines are provided for evaluating convicted offender DNA typing profiles obtained by STR analysis using PowerPlex Fusion 6C, an AB Genetic Analyzer, and GeneMapper *ID-X*.

DATA ANALYSIS WORKFLOW

Creating a Project in *ID-X*

1. In the new Project window, select Add Samples to Project. Under the Files tab, navigate to the desired run folder, then click Add to List followed by Add.
2. In the Project window, select the desired view from the Table Settings drop-down menu.
3. In the Samples tab, ensure that the correct Sample Type, Analysis Method, Panel, and Size Standard are selected. Select a custom control, if applicable.
4. Click the Analyze (green arrow) button and complete the fields in the Save Project dialog box. Click OK to save the project.

NOTE: Should any of the samples be inadvertently misnamed, the above steps will be completed prior to re-naming the sample in *ID-X*. This ensures that the edit will be documented in the sample's Info tab in the software.

5. In the navigation pane, expand the run folder, select the Raw Data tab, and scroll through each sample. Verify the presence of primer peaks in all samples. The raw data may also aid in the evaluation of anomalies and causes of poor sizing quality.

Quality Assessment & Documentation in *ID-X*

GeneMapper *ID-X* analyzes CE data and assists the analyst in evaluating sample quality through a system of process quality values (PQVs). These PQVs may be either sample- or marker-level and are associated with color-coded flags that are triggered when a sample does not meet the quality requirements defined in the software. The PQV results are displayed as:

- Pass (green square)
- Check (yellow triangle)
- Low Quality (red octagon)

Evaluate the quality flags to help determine whether a sample may be accepted or requires additional testing. Once deemed acceptable, a system of peak edits and marker-level Genotype Quality (GQ) flag overrides is used to document analyst review of the sample.

Peak Edits: Select the peak, then right-click and select Rename Allele Label. Identify artifacts using the pre-defined list or the Custom Artifact Label; artifact labels will appear in pink boxes in the electropherogram. True alleles requiring re-naming (e.g. microvariants) will be edited using the Custom Allele Label. Each edit is tracked in the software.

Genotype Quality PQVs: Once all edits are complete for a locus, select the marker in the Genotypes table, right-click, then select Yes in the dialog box to override the GQ. The GQ PQV will turn into a passing flag and all other PQVs for the marker will turn gray and retain their original shape. When all GQs have been overridden for a sample, the user will be prompted to override the Composite Genotype Quality (CGQ). Select No.

Samples that do not pass will remain in the run folder to aid technical review, but do not need to be analyzed using peak edits and GQ overrides.

DATA EVALUATION & INTERPRETATION

All samples and controls shall be examined, regardless of the PQV flags.

Raw Data

Examine the raw data. Typically, the primer peak should rise and fall abruptly and the baseline should be flat and smooth. If the primer peak trails excessively or the baseline is elevated, negatively affecting genotyping quality, the sample should be re-injected.

WEN Internal Lane Standard

The size standard data in each sample will be examined. If the Sizing Quality (SQ) PQV flag is yellow or red, the data should be inspected to confirm that all of the peaks are present and correctly labeled. All size standard peaks must be above the analytical threshold and within the analysis range. Other quality issues, such as broad peaks, may be accepted at analyst discretion if genotyping quality is not adversely affected.

Allelic Ladder

An appropriate allelic ladder will be included in every run folder. All ladders will be reviewed and any flagged ladders will be evaluated to determine if the failure was due to a factor that may have affected the corresponding injection. *ID-X* does not use flagged ladders for data analysis.

Controls

All controls will be reviewed by the analyst. It may be appropriate for the analyst to override PQV flags depending on the type of quality issue.

- Reagent Blank: No forensically significant PCR product should be detected in the reagent blank. The reagent blank shall be amplified using the same primers, instrument model, and concentration conditions as required by the sample(s) containing the least amount of DNA. The amplified reagent blank shall be typed using the same instrument model, injection conditions, and most sensitive volume conditions of the extraction set.
- Positive Amplification Control (+C): Confirm that the genotype is correct and no forensically significant contamination is present. The +C is not subject to confirmation of peak height imbalances. When multiple +Cs exist for an amplification, only one must pass.
- Negative Amplification Control (-C): No forensically significant PCR product should be detected in the -C. It must be loaded and injected under at least the same sensitivity conditions as the corresponding +C. For direct amplifications, the -C is a combined reagent blank and negative amplification control; it must be handled like a reagent blank.
- Internal Laboratory Control (Custom Control): When used, the internal laboratory control should genotype correctly and be free of contamination. Confirmation of peak height imbalances is not required.

Samples

Analysts will evaluate all samples based on the guidelines that follow. Workflow is at analyst discretion, although it is recommended to review samples that have met all analysis thresholds first, followed by samples with triggered PQV flags. The Analysis Summary tab in *ID-X* can be used to navigate easily to these two groups of samples.

True Alleles

True alleles are sharp, distinct peaks that are equal to or greater than the analytical threshold and typically have an estimated size between 60 and 500 bp. Peaks that are saturated (off-scale) or broad, but genotyping is not adversely affected, may be accepted by the analyst, in conjunction with the technical reviewer.

Determination of a genotype is aided by the use of thresholds established during validation.

Analytical Threshold: 120 rfu

The analytical threshold (detection threshold) is the level at which a peak can reliably be detected above background fluorescence.

Stochastic Threshold: 500 rfu

The stochastic threshold is the minimum peak height above which there is a high probability that both alleles of a heterozygous locus will be detected.

Heterozygous alleles are expected to appear as two peaks at least as high as the analytical threshold and have a peak height ratio of 60% or greater. Peak height ratios less than 60% will be confirmed by re-injection (Amelogenin is exempt from this requirement). Should the peak height imbalance persist, the analyst, in conjunction with the technical reviewer, may accept the sample. Otherwise, the sample shall be re-amplified.

Homozygous alleles are expected to appear as single peaks at least as high as the stochastic threshold.

For Y STR loci, alleles are expected to appear as single peaks at least as high as the analytical threshold.

Variant Alleles

If an off ladder (microvariant) allele, tri-allele, or duplication at a Y STR locus is observed, the NIST STRBase website <http://www.cstl.nist.gov/strbase/> will be checked to see if the variant has been previously reported using a CE platform and Fusion 6C. Novel variants will be reported to STRBase after proper confirmation and technical review.

Off ladder alleles may be accepted without additional analysis if they have been previously reported at STRBase and the sizing quality of the sample is passing. The sample will be re-injected for confirmation if sizing quality is not acceptable. If an off ladder allele has not been previously reported, the sample will be re-injected, at a minimum, for confirmation.

Occasionally a one base pair difference between two alleles cannot be resolved using typical methods like re-injection or re-amplification. In such cases, the Polynomial Degree and Peak Window Size in *ID-X* may be adjusted. This will be documented in the case file and the modified analysis method will be included in the electronic data.

The following table will be used as a guide for allele designations outside of the ladder. For off ladder alleles above or below the limits in the table and given an OL designation in *ID-X*, the applicable < or > will be used as the allele designation. For manual CODIS tables, the upper and lower values in the table must be used.

LOCUS	Lower	Upper
D3S1358	<12	>19
D1S1656	<9	>20.3
D2S441	<9	>16
D10S1248	<8	>18
D13S317	<8	>15
Penta E	<5	>24
D16S539	<5	>15
D18S51	<9	>26
D2S1338	<15	>28
CSF1PO	<6	>15
Penta D	<2.2	>17
TH01	<5	>10
vWA	<11	>21
D21S11	<24.2	>38
D7S820	<6	>14
D5S818	<7	>16
TPOX	<6	>13
D8S1179	<8	>18
D12S391	<14	>27
D19S433	<9	>17.2
SE33	<4.2	>37
D22S1045	<8	>19
DYS391	<8	>13
FGA	<18	>30
DYS576	<11	>23
DYS570	<10	>25

Tri-alleles and Y STR duplications that have been previously reported on STRBase will be re-injected for confirmation. If not reported, the sample will be re-injected, at a minimum, for confirmation.

Abnormal sex chromosome variants (e.g. XXY) do not need to be re-tested.

PCR and Genetic Analyzer Artifacts

If an artifact can be confidently characterized as one of the following anomalies, the sample does not need to be re-tested. If there is any uncertainty about an anomalous peak or it may interfere with possible alleles, the sample should be re-run to ensure quality and accuracy of the allele calls.

An artifact is acceptable if it meets the criteria defined below, is appropriately labeled by the analyst in *ID-X*, and the technical reviewer agrees that the artifact has been accurately characterized and documented, and is not due to contamination or another problem.

▪ **Spikes**

Spikes have a general appearance of being needle-like and are randomly seen in samples. Spikes typically disappear with re-injection. Commonly, spikes are observed in most or all of the dye colors and generally at the same base pair location. Occasionally, single color spikes may also be observed. The heights of spikes usually vary between dye colors.

ID-X may automatically flag spikes within the analysis range. Spikes may also be flagged as something other than a spike and other artifacts, such as pull-up, may be labeled as spikes. Evaluate the artifact and PQV flags, and appropriately label the artifact.

▪ **Pull-up**

Pull-up artifacts are peaks that show up at the same base pair size or close (i.e. about a base equivalent on either side of the allele), but in another color. Pull-up usually occurs under peaks with excessive signal and may also occur between strong peaks.

▪ **Stutter**

Stutter is an artifact that occurs during PCR due to strand slippage and is a well-characterized amplification phenomenon. Stutter peaks are generally one repeat unit less than or greater than the allele, although n-2 and n-8 stutter peaks may also be observed. Established stutter percentages aid in discriminating between stutter and other potential alleles (e.g. weak heterozygous or tri-allele peaks, contamination). Stutter filters are set in *ID-X* using the values below.

Locus	n-4 / n-3 / n-5 % Stutter	n-2 % Stutter	n+4 / n+3 / n+5 % Stutter
Amelogenin	--	--	--
D3S1358	13.5	--	1.7
D1S1656	14.3	3.6	2.3
D2S441	9.0	--	1.8
D10S1248	13.0	--	1.3
D13S317	10.3	--	2.2
Penta E	7.2	--	1.9
D16S539	12.0	--	3.0
D18S51	14.6	--	2.8
D2S1338	13.6	--	2.2
CSF1PO	11.1	--	3.7
Penta D	4.5	--	3.7
TH01	4.8	--	1.5
vWA	14.4	--	2.7
D21S11	12.7	--	2.8
D7S820	9.7	--	1.8
D5S818	11.0	--	2.3
TPOX	5.4	--	1.1
D8S1179	11.8	--	3.4
D12S391	17.4	--	2.7
D19S433	12.1	1.4	2.6
SE33	16.1	6.6	3.3
D22S1045	16.8	--	9.0
DYS391	9.4	--	2.0
FGA	12.4	1.2	2.8
DYS576	12.5	--	3.4
DYS570	13.0	--	2.4

When stutter peaks exceed the threshold by a minimal amount, it is acceptable for the analyst to re-label the peak. Larger deviations should be evaluated closely and may require re-amplification or extraction to determine if they are actual stutter peaks.

▪ **DNA-Dependent Artifacts**

The following DNA-dependent artifacts have been identified by the manufacturer of Fusion 6C and may be appropriate for the analyst to re-label after careful evaluation.

Locus	Artifact Size
Amelogenin	n-1
D1S1656	n-1, n-2
D13S317	n-2
D18S51	n-2
vWA	n-2, elevated baseline
D7S820	n-2
D5S818	n-2
D19S433	n-2
SE33	n-2
DYS391	n-1
FGA	n-1, n-2

▪ **Shoulders**

Shoulders flank the sides (one or both) of an allele. When shoulders are given a designation in *ID-X* and are 10% or less of the parent allele, they may be re-labeled. Otherwise, the sample should be re-run at a shorter injection time, diluted with TE, or re-amplified.

▪ **Dye Blobs**

Raised baseline areas may represent dye blobs, which often occur around the same base pair location in multiple samples of a run.

Contamination

In instances of a suspected contamination event where the questioned PCR product is above the analytical threshold, troubleshooting will be conducted to determine where in the laboratory process it may have occurred. The contamination flow charts in the DNA Quality Assurance Manual should be consulted to assist with troubleshooting. Re-testing of samples or controls with suspected contamination below the analytical threshold is not required.

REPORTING PROFILES & CODIS DATABASE

Profiles deemed appropriate to enter into the CODIS database must be peer reviewed prior to entry. In cases where a full profile could not be generated, the partial profile may still be eligible for SDIS. Refer to the WSP CLD CODIS SOP Manual for further guidelines on entering convicted offender sample profiles in SDIS and NDIS.

Profiles eligible for CODIS are reported in either electronic or manual tables; electronic tables should be used whenever possible. Manual tables may also be used for reporting profiles for hit confirmations.

Electronic CODIS Tables

1. In *ID-X*, select the appropriate Specimen Category for each sample being exported.
2. Select the desired run folder in the Navigation Pane, then select File > Export Table for CODIS.
3. Ensure that the correct source and destination laboratories are selected and that the file type is CMF 3.2 (.xml).
4. Enter the file name, which should be the same as the run folder name, and click Export.

Manual Tables

1. Make an allele table using Microsoft Excel, being sure to include all loci and reported alleles. This information can be exported from *ID-X* rather than transcribing.
2. For partial profiles, single alleles above the analytical threshold and below the stochastic threshold (or any inconclusive genotype above the analytical threshold) will be in parentheses.

These alleles will not be entered into CODIS, but may be used for exclusion or evaluating matches.

Composite Profiles

In instances where a sample has been consumed and a full profile was not developed, but a more complete or full profile may be constructed by considering results from multiple tests of the sample, a composite profile may be generated. For a locus to be considered for use in the composite profile, it must meet the reporting requirements described in this manual.

A manual table will be generated for reporting composite profiles. All run folders used for constructing the profile will be indicated on the Run Folder Review sheet.

GUIDELINES FOR EVALUATING CONVICTED OFFENDER DNA TYPING PROFILES USING GENEMAPPER® ID-X V1.4 AS AN EXPERT SYSTEM

This procedure shall be followed to optimize workflow and expedite data analysis when using GeneMapper ID-X as an Expert System.

EVALUATING LADDERS AND CONTROLS

1. From the Analysis Summary tab:
 - a. Any passing allelic ladders do not have to be evaluated.
 - b. Failed allelic ladders must be evaluated by an analyst to verify the failure wasn't due to a factor that affected the corresponding injection.
 - c. In the case of more than one of the same kind of control (i.e. amplification positive controls that were amplified at the same time on the same thermal cycler), only one iteration need pass. If there is not a positive control under the "All Thresholds Met" column, view the positive control(s) under the "One or more thresholds not met" column and evaluate why they did not pass. Follow the guidelines already described [in this manual](#) for evaluation of positive controls.
 - d. Select the negative controls and verify no trace contamination exists.

EVALUATING SAMPLES

An analyst is only required to view samples under the "One or more thresholds not met" column. To maximize workflow, an analyst should not view samples under the "All Thresholds Met" column for routine data analysis.

1. If the project contains samples to be uploaded to CODIS, select the number of the samples under the "All Thresholds Met" column under the "Analysis Summary" tab. Change the "Export Type" of all samples in this column to the appropriate designation. Otherwise, proceed to the next step.
2. From the "Analysis Summary" tab, select the number of the samples under the "One or more thresholds not met" column.
3. Highlight a sample and display the plot. Use a plot setting that allows for the "Quality Value Details" box to be visible.
4. Select the first locus with a flag and evaluate the description of the flag in the "Quality Value Details" box.
5. If the flag is something an analyst can override:
 - a. Make the appropriate change and "Override the Genotype Quality" at that marker. Appropriate overrides are discussed previously [in this manual](#).
 - b. Continue evaluating other flags in the sample; override those that are appropriate to override.
 - c. When prompted to override the "Composite Genotype Quality", select No.

- d. Evaluate the remaining non-flagged loci to ensure there are no anomalies or evidence of contamination and that the analyst agrees with the allele calls.
6. If a flag is not appropriate to be overridden, continued evaluation of the sample is unnecessary except in terms of troubleshooting what steps (in the laboratory or otherwise) to take next with that sample.
7. If more than two true allele peaks occur at multiple loci (two or more), the profile should be considered as having originated from more than one individual. Occasionally, three alleles may be detected at any one locus. If there is no other indication of more than one source for the DNA profile, it is permissible to assume a single source and accept the genotype result for CODIS after the appropriate confirmations are made.
 - a. If contamination is suspected, refer to [Appendix II of the DNA Quality Assurance Manual](#) for the appropriate troubleshooting steps.
8. If an analyst overrode the flags to pass the sample, change the Export Type to the appropriate designation.

Repeat the process for the remaining samples in the "One or more thresholds not met" column.

CODIS TABLES

If manual CODIS tables need to be prepared, make those prior to technical review, print the table and include it in the case file.

For electronic CODIS tables, export any samples for CODIS after the technical reviewer has evaluated the run folder. When exporting samples for CODIS tables, select an unfiltered view of the project to ensure inclusion of all exportable samples.

CODIS CASE FILE CONTENT

A CODIS Case File consists of all documentation related to CODIS Casework including hard copy and electronic data and/or worksheets. A CODIS Case File will contain, at minimum, the following items filled out with sufficient information as to appropriately record all of the pertinent data.

1. Extraction Worksheet (for any manual extractions)
2. 7500 Load Sheet (if applicable)
3. Amplification Worksheet (for any manual amplification set-ups)
4. 96 Well Worksheet (for any 96-well plate sets)
5. Run Folder Review Sheet
6. CE Loading & Reagent Worksheet (if applicable)
7. Manual CODIS Table (when appropriate)
8. CODIS Table (when appropriate)
9. PDF of the puncher work list (when appropriate)
10. Quantifiler™ Projects
11. Raw Data folder(s) containing only those samples and controls related to a particular set
12. GeneMapper® ID-X project (exported with analysis settings)
13. Peer Review Checklist (to be added by the technical reviewer)

MANUAL CODIS CASE FILE REVIEW WITH GENEMAPPER® ID-X V1.4

There are two fundamental steps to the review of CODIS case files which are to be performed by a qualified CODIS scientist:

1. Review of the electronic data.
2. Review of the associated paperwork from extraction through analysis to verify documentation is accurate.

It is the responsibility of the reviewer to ensure the analyzed data is as accurate and complete as possible. At minimum the following will be done:

1. Verify that all electronic documents and files are present.
2. Check the GeneMapper® ID-X results.
 - a. The ladders and quality control samples will be evaluated by the reviewer to ensure they meet the requirements as described in the Guidelines for Evaluating DNA Typing Profiles using GMID-X section of this manual.
 - b. The reviewer must look at all samples (to include the ILS) reported from a run folder.
 - i. For samples where a flag was overridden or an edit was made, the reviewer must evaluate the changes.
 - ii. When an edit has been made, the reviewer must ensure the original analyst documented the change appropriately in the GMID-X software.
 - iii. If the reviewer agrees with the edited profile and/or the analyst's assessment to override flags, the reviewer indicates this by overriding the Composite Genotype Quality (CGQ) of the sample.
 - iv. For passing samples, the reviewer will evaluate the profiles for allele designations and that the sample is free from extraneous peaks.
 - c. All samples shall meet the requirements outlined in the Guidelines for Evaluating DNA Typing Profiles using GMID-X section of this manual. In instances where they do not, but there is cause to report the sample, the CODIS supervisor or DNA Technical Leader may sign off on those samples if it is viewed as appropriate. This must be reflected in the case file.
 - d. The CODIS tables will be evaluated to ensure the samples are reported from the correct run folder, the table is the correct export version, each sample has the appropriate Specimen Category, and that the analyst's CODIS username is correct.
3. Review the paper worksheets.
 - a. Check that all information is filled out on the worksheets and that the appropriate worksheets are present.
 - b. Check that the case numbers of the convicted offender samples are accurate across all worksheets.

- c. Verify that a set number, date, and initials are on each page.
- d. Ensure that all strikeouts and additions are initialed.

For a more detailed list of what is required for the technical review of a CODIS Case File, see the CODIS Laboratory Peer Review Checklist.

CODIS CASE FILE TECHNICAL REVIEW WITH GENEMAPPER® ID-X V1.4 AS AN EXPERT SYSTEM

REVIEW OF LADDERS

Ladders do not have to be evaluated by a technical reviewer.

REVIEW OF CONTROLS

The ILS of a control under the “All Thresholds Met” column does not have to be reviewed by the technical reviewer.

From the Analysis Summary tab:

1. Check the negative controls under the “All Thresholds Met” column for trace contamination.
2. Do not look at the other controls under the “All Thresholds Met” column.
3. If any controls are under the “One or more thresholds not met” column, click on the number under the total number of controls in that column.
4. Look at any control reported from the project to verify analyst overrides are appropriate.
5. For any control with allele edits, use the “View Allele Edits” plot setting to verify artifacts/alleles are documented appropriately.
6. If in agreement with the control overrides/edits, override the CGQ of the control.
7. For projects with more than one positive amplification control from the same amplification, only one of those controls must pass.

REVIEW OF SAMPLES

The ILS of a sample under the “All Thresholds Met” column does not have to be reviewed by the technical reviewer.

1. Do not view the genotypes of samples under the “All Thresholds Met” column.
2. Click on the number under the “One or more thresholds not met” column.
3. Look at any sample reported from the project to verify analyst overrides are appropriate.
4. For any sample with allele edits, use the “View Allele Edits” plot setting to verify artifacts/alleles are documented appropriately.
5. If in agreement with the sample overrides/edits, override the CGQ of the sample.
6. Continue for all reported samples from the run folder.

Continue reviewing the case file and other projects, if applicable.

For confirmation runs, it is appropriate to review the original iteration of the sample to verify an appropriate confirmation was made.

Review the electronic CODIS tables to ensure the appropriate samples are reported from the correct run folder with the correct Export Type, export version, and analyst CODIS username.

For additional details on the requirements of technical review, see the [CODIS Laboratory Peer Review Checklist](#).

TEMPGENIUS™ WIRELESS DATA ACQUISITION & MONITORING SYSTEM

The TempGenius system is used to wirelessly record the temperature for refrigerators, freezers, and ovens in the lab. Rather than recording the temperature of a unit at one time of day, TempGenius constantly monitors and records the temperature. If a unit falls outside of the desired temperature range, TempGenius will alert the user and record the event. This allows for more effective monitoring of refrigerators, freezers, and ovens.

GENERAL INSTALLATION AND OPERATION

Refer to the TempGenius “Installation Guide And Operating Manual”.

ALERTS AND RESOLUTIONS

TempGenius automatically monitors temperatures and sends alerts to the assigned user(s) and TempGenius Dashboard. In addition to viewing temperatures on the Dashboard, real-time temperatures may be viewed remotely by accessing the IP address of the TempGenius Data Server. Refer to the TempGenius Installation Guide and Operating Manual for instructions on alerts and alert resolutions.

TROUBLESHOOTING ALERTS

Use the following table as a guide for troubleshooting alerts.

Alert Type	Critical Unit?	Action	Address Off-hours?
Level 1	No	If more than two days in a row, verify reading against a NIST thermometer. If NIST is within allowable range, contact the TempGenius rep, otherwise troubleshoot until a solution is reached.	No
Level 1	Yes	If more than two days in a row, verify temperature reading against a NIST thermometer. If NIST is within the allowable range, contact the TempGenius rep, otherwise move contents of unit to another location until a solution can be reached.	No
Level 2	No	Verify reading against a NIST thermometer. If NIST is within allowable range, contact the TempGenius rep, otherwise troubleshoot until a solution is reached.	No
Level 2	Yes	Verify the temperature reading against a NIST thermometer. If NIST is within the allowable range, contact the TempGenius rep, otherwise move contents of unit to another location until a solution can be reached.	Yes, if a volunteer is unavailable, notify CODIS Manager
Level 3	No	Verify the temperature reading against a NIST thermometer. If NIST is within the allowable range, contact the TempGenius rep, otherwise troubleshoot as needed.	No
Level 3	Yes	Verify the temperature reading against a NIST thermometer. If NIST is within the allowable range, contact the TempGenius rep, otherwise move contents of unit to another location until a solution can be reached.	Yes, if a volunteer is unavailable, notify CODIS Manager

REAGENT PREPARATION

Use reagent grade chemicals unless otherwise noted. Pre-made, reagent grade chemicals can be substituted for many of the recipes below. Prepare all solutions using deionized water (dH₂O). Wear gloves and follow safety recommendations provided by manufacturer for handling chemicals. Comply with any and all laws, regulations, or orders with respect to the disposal of any hazardous or toxic chemical, material, substance or waste. Reagents that are used solely for the preparation of other reagents do not need to be sterilized before making up the new reagent if that reagent will be sterilized. Store all reagents at room temperature unless otherwise noted. Reagents will be labeled with the reagent name, lot number, expiration date (if any), storage requirements (if not at room temperature), preparer's initials, and date of preparation. Reagents are prepared as follows:

0.5M EDTA (1 L)

Warning: EDTA is an irritant. Wear lab coat, gloves, mask, and protective eyewear when handling EDTA.

Warning: NaOH is corrosive and toxic. Wear lab coat, gloves, and protective eyewear. Use caution when handling. NaOH, when combined with water, results in an exothermic reaction. Dissolve the NaOH pellets in the water gradually with 3 or 4 additions. Cover and allow the NaOH to dissolve completely and cool between each addition.

Add 186.1 g of disodium ethylenediaminetetraacetic acid dihydrate (Na₂EDTA · 2H₂O) to 800 ml of dH₂O. Stir vigorously with a magnetic stirrer. To dissolve the EDTA powder, adjust the pH to 8.0 (+/- 0.2) by adding approximately 20 g of NaOH pellets. Check the pH and add 5N or 10N NaOH solution if needed. Adjust volume to 1 liter with dH₂O and mix thoroughly. The solution should be autoclaved or filtered through a sterile 0.2 µm filter.

TE BUFFER [10 MM TRIS-HCL, 0.1 MM EDTA, PH 8.0 (1 L)]

Add 10 ml of 1 M Tris-HCl, pH 8.0 and 0.2 ml of 0.5 M EDTA to 990 ml dH₂O and mix thoroughly. Autoclave.

1 M TRIS-HCL, PH 8.0 (1 L)

Dissolve 98 g Tris-HCl and 46 g Tris base in 800 ml of dH₂O. Check that the pH of the solution is 8.0 (+/- 0.2). Adjust final volume to 1 liter with dH₂O and mix thoroughly. Sterilize by autoclaving.

1 M DITHIOTHREITOL (DTT - 10 ML)

Add 1.542 g of DTT to 10 ml dH₂O. Aliquot 1 ml of solution to ten tubes. Store frozen.

LYSIS BUFFER + DTT

Prepare Lysis Buffer + DTT by mixing 1 ml of 1.0 M DTT for every 100 ml DNA IQ™ Lysis Buffer. Mix by inverting several times. Store at room temperature.

1X WASH BUFFER

Prepare 1X Wash Buffer by mixing one part 95-100% ethanol, one part isopropyl alcohol, and two parts DNA IQ™ 2X Wash Buffer. Mix by inverting several times. Store at room temperature.

QUANTIFILER™ QUANTIFICATION STANDARDS

DNA quantitation standards are critical for accurate analysis of run data. Any mistakes or inaccuracies in making the dilutions directly affect the quality of the results. Great care shall be taken when measuring and mixing dilutions.

1. Prepare the DNA Quantitation Standards using G147A (Promega catalog #G1471) genomic DNA (SRM 2372 traceable)
2. Label eight disposable tubes to be used for the dilution series.
3. Prepare the TE with glycogen by adding 1µL glycogen to 1 ml of TE. Vortex.
4. Dispense the required amount of TE/glycogen to each tube. (See note* on next page to adjust for lot to lot variation in stock DNA standard.)
5. Prepare the standards using the following table.
 - a. For standard 1:
 - i. Briefly vortex the G147A Human DNA Standard and pulse spin.
 - ii. Using a new pipette tip, add the calculated amount of G147A Human DNA Standard to the tube for Standard 1.
 - iii. Vortex and pulse spin the dilution
 - b. For standards 2 through 8:
 - i. Using a new pipette tip, add the calculated amount of the previously prepared standard to the tube for the next standard.
 - ii. Vortex and pulse spin the standard.
 - iii. Repeat steps d.2.i and d.2.ii until the dilution series is completed.
6. The prepared Human DNA Standards may be stored for up to 4 weeks at 2 to 8°C.

Standard	Concentration (ng/µL)	Example Amounts	Minimum Amounts [†]	Dilution Factor
Std. 1	50.000	50 µL [200 ng/µL stock] + 150 µL TE/glycogen buffer	10 µL [200 ng/µL stock] + 30 µL TE/glycogen buffer	4X
Std. 2	16.700	50 µL [Std. 1] + 100 µL TE/glycogen buffer	10 µL [Std. 1] + 20 µL TE/glycogen buffer	3X
Std. 3	5.560	50 µL [Std. 2] + 100 µL TE/glycogen buffer	10 µL [Std. 2] + 20 µL TE/glycogen buffer	3X
Std. 4	1.850	50 µL [Std. 3] + 100 µL TE/glycogen buffer	10 µL [Std. 3] + 20 µL TE/glycogen buffer	3X
Std. 5	0.620	50 µL [Std. 4] + 100 µL TE/glycogen buffer	10 µL [Std. 4] + 20 µL TE/glycogen buffer	3X
Std. 6	0.210	50 µL [Std. 5] + 100 µL TE/glycogen buffer	10 µL [Std. 5] + 20 µL TE/glycogen buffer	3X
Std. 7	0.068	50 µL [Std. 6] + 100 µL TE/glycogen buffer	10 µL [Std. 6] + 20 µL TE/glycogen buffer	3X
Std. 8	0.023	50 µL [Std. 7] + 100 µL TE/glycogen buffer	10 µL [Std. 7] + 20 µL TE/glycogen buffer	3X

[†]To ensure the accuracy of pipetting, the minimum input volume of DNA for dilutions is 10 µL.

***NOTE: How to adjust for lot to lot variation in the Human DNA Standard:**

Use the SRM 2372 kit from NIST to determine the DNA concentration of each new lot of G147A DNA standard. For step 1 of preparing the Quantitation Standard sets, dilute the stock according to the actual concentration obtained for the new lot of DNA. The rest of the dilution steps remain

the same.

DISPENSING HI-DI FORMAMIDE

Hi-Di Formamide is harmful if swallowed, inhaled, or absorbed through the skin. See MSDS for full disclaimer. When aliquoting Hi-Di Formamide, work in a fume hood with gloves and lab coat. Remove the Hi-Di bottle from the freezer and allow it to come to room temperature until contents have thawed. Aliquot 1 mL Hi-Di into 1.5 mL autoclaved tubes. Label each tube with the lot, date of dispensing, and initials. Freeze immediately.

REAGENT EXPIRATION DATES

The following table lists expiration dates for reagents used in the CODIS Laboratory. Reagents that are tested before each use do not require an expiration date. Should a stock solution purchased from a distributor be used rather than being made in the lab, the expiration date on the bottle overrides any expiration date listed here.

Commercially prepared critical reagents that do not have an expiration date assigned by the manufacturer will be assigned one by the laboratory based on performance testing and review of scientific literature or up to three years from receipt.

Non-critical commercially prepared reagents that do not have an expiration date assigned by the manufacturer do not need one assigned by the laboratory provided they are stored and maintained appropriately. The integrity of these reagents, as with any reagent, will be monitored at the time of use.

Reagent Expiration Dates

DTT	3 years (stored frozen)
0.5 M EDTA	3 years
TE Buffer	2 years
1M Tris-HCl	3 years
Lysis Buffer + DTT	1 month
1X Wash Buffer	1 year
QF Standards	4 weeks if in glycogen

TEMPGENIUS™ WIRELESS DATA ACQUISITION & MONITORING SYSTEM MAINTENANCE

SEMIANNUALLY

A NIST probe reading for each monitored unit will be recorded semiannually, to ensure the accuracy of the temperature readings recorded by TempGenius. If the difference between the TempGenius reading and the NIST reading is outside the allowable range, another reading will be taken at a different time. In the event that the TempGenius reading is once more outside the range, contact the manufacturer.

For Daylight Savings time changes, a “write setup” from the Pointware program to the data receiver must be performed to sync the computer and device time. Before performing the write setup, ensure that the historical data collections are current. This can be done by making sure the time of “last contact” listed in Pointware (under the Point Managers tab) is the same as the current time on the computer.

Once the historical data are current, perform the write setup in Pointware via the Point Managers tab.

1. Right-click on the appropriate Point Manager name
2. Select “Push”
3. Select “Write Setup”
4. Select “Yes”.

ANNUALLY

Calibration of sensors – see the TempGenius “Installation Guide and Operating Manual”

The manufacturer defines the annual calibration as the adjustment (or replacement) of a sensor due to temperature variability against a NIST probe. The average offset for a sensor will be determined based on the total readings for the year. The laboratory may choose to make an adjustment if the offset between the sensor and NIST probe is +/-1°C. If the average offset of a sensor is more/less than +/- 4°C, an adjustment is mandatory and the manufacturer will be contacted.

If any adjustment is made to the sensor, a reading with A NIST probe will be taken the following day to verify the adjustment. All adjustments will be documented on the TempGenius maintenance sheet.

EVERY 30 MONTHS OR AS NEEDED

Replace the batteries in the sensors by unscrewing the transmitter cover.

UV IRRADIATOR OPERATING INSTRUCTIONS

OPERATION

1. Turn the power on. The digital display should read 0000 or 0; if it does not, press reset.
2. Place the consumables in an orientation to allow maximum exposure.
3. Select the Energy button. The number on the display represents microjoules/cm² x 100.
 - For autoclaved consumables, set the microjoule level to 500,000 µJ/cm².
 - For non-autoclavable consumables, set the microjoule level to 999,900 µJ/cm²
4. If an error is made in entering, press the reset button to clear the display.
5. Press start to begin irradiation.

MAINTENANCE

1. Clean the floor of the UV Irradiator with a soft cloth and nonabrasive cleaner once a week.
2. Change bulbs as needed.

BSD600 DUET PUNCHER MAINTENANCE

BEFORE EACH USE

1. Clean the outside of the puncher, the chute, and the punch mechanism. See user manual for detailed instructions.
2. Empty the waste punch container if needed.
3. Ensure the sponges in the humidifier bottles are moist.

ANNUAL

On a yearly basis, the BSD600 Duet puncher will be serviced by a qualified technician.

PERFORMANCE CHECKS

A performance check will be done following repair or routine maintenance by a qualified technician. The check must be completed prior to use for convicted offender sample analysis and will address the reliability of the instrument.

ABI 7500 REAL TIME PCR SYSTEM MAINTENANCE

MONTHLY

1. Check the lamp status
2. Although background calibration is valid for six months in the 7500 Software, the block shall be checked monthly for contamination by performing a background calibration in the Instrument Maintenance Manager.
3. Perform a Function Test

SEMI ANNUALLY

1. Check the Lamp Status
2. Perform all calibrations as prompted by the 7500 Software Instrument Maintenance Manager. The RNase P run does not need to be performed.
3. Perform a Function Test

ANNUALLY

Preventive maintenance, including all semi-annual maintenance tasks, will be performed by a qualified technician

PERFORMANCE CHECKS

A performance check will be done on an instrument if it has been moved or following repair or routine maintenance done by a qualified technician. The check must be completed prior to use for convicted offender sample analysis and consists of:

1. ROI, background, optical, and Pure Dye calibrations (may be performed by the technician)
2. A quantification plate containing previously quantified samples.

9700 THERMAL CYCLER MAINTENANCE

BIANNUAL

1. Clean the sample wells and heated cover
2. Run the Calibration Verification test
3. Run the Temperature Non-Uniformity test

AS NEEDED

Send to a qualified technician for external calibration or repair.

PERFORMANCE CHECK

A performance check will be done following repair or maintenance by a qualified technician. The check must be completed prior to use for convicted offender sample analysis and will consist of the Calibration Verification and Temperature Non-Uniformity tests plus positive and negative amp controls.

3500XL GENETIC ANALYZER MAINTENANCE

SPATIAL CALIBRATION

A spatial calibration must be performed every time a capillary array is removed or replaced, the detector door is opened, or the instrument is moved.

1. In the Maintenance navigation pane, select 'Spatial Calibration'.
2. Select either 'Fill' or 'No Fill', then start the calibration.
3. Evaluate the calibration results to ensure that there is:
 - a. A sharp peak for each capillary (small shoulders are acceptable).
 - b. A '+' marker at the apex of each peak and no off-apex markers.
 - c. An even peak profile (all peak about the same height).
4. Accept the results if the criteria above are met. Refer to the 3500xL User Guide if the calibration fails.

SPECTRAL CALIBRATION

Perform a spectral calibration when a new dye set/polymer combination is used, if a new capillary array is installed, after any service on the optics, or if there is an increase in pull-up observed.

At first use, thaw the 6C Matrix Mix and Matrix Mix Dilution Buffer completely. After first use, store the reagents at 2 – 10°C.

1. Pre-heat the oven to 60°C, ideally 30 minutes before the first injection.
2. Vortex the 6C Matrix Mix for 10 – 15 seconds. Add 10 µL of 6C Matrix Mix to one tube of Matrix Dilution Buffer. Vortex for 10 – 15 seconds. The diluted 6C Matrix Mix can be stored for up to 1 week at 2 – 10°C.
3. Add 10 µL of the diluted 6C Matrix Mix prepared in step 2 to 500 µL of Hi-Di Formamide. Vortex for 10 – 15 seconds.
4. Add 15 µL of the 6C Matrix Mix with formamide prepared in step 3 to each well of columns 1 – 3 of a 96-well plate. Cover with a plate septum and centrifuge briefly to remove bubbles. DO NOT heat denature.
5. Place the assembled plate onto the instrument. Select 'Spectral Calibration' in the Maintenance navigation pane. Select 'Matrix' for the chemistry standard and 'Promega J6' for the Dye Set.
 - a. The software will automatically run up to 3 injections to obtain a passing calibration.
 - b. Not allowing capillary borrowing generally produces better results, but the software may be configured to allow up to 3 borrowing events.
6. Start the calibration.
7. Evaluate the results and choose 'Accept' or 'Reject'. Refer to the 3500xL User Guide for detailed instructions on evaluating calibration data.

The 3500 Series Software employs wizards to assist the user in performing maintenance procedures. Access the appropriate wizard from the Maintenance Wizard screen and follow the prompts to complete any required tasks. The Maintenance calendar on the Dashboard may also be used to remind users when task are due.

BEFORE EACH RUN

- Check for any bubbles in the pump block and lines. If present, use the Remove Bubble Wizard.
- Check consumables on the Dashboard. Replace polymer and buffers if needed.

WEEKLY

- Restart the instrument and computer.
- Clean the anode buffer container valve pin.

MONTHLY

- Run the Wash Pump and Channels wizard and replace the polymer.
- Flush the pump trap.
- Replace the cathode buffer container septa.
- Clean the autosampler and drip tray.

QUARTERLY

- Run a sensitivity check:
 1. Quantify 2800M DNA. This DNA should be kept in Teflon microcentrifuge tubes when stored for extended use.
 2. Amplify 1.0 ng of quantified 2800M DNA.
 3. Load such that each capillary injects amplicon (one well reserved for ladder) and inject for 24 seconds. Rotate the ladder position so that data is generated for each capillary over the course of the year.
 4. Ensure that the average heterozygous peak height (homozygous peaks are halved) for each capillary is $\geq 50\%$ of the previous run. Drops in sensitivity outside this range require troubleshooting.

ANNUALLY

- Preventive maintenance will be performed by a qualified technician.

AS NEEDED

- Replace the capillary array using the Install Capillary Array wizard when poor precision or allele calling is noted, or when resolution or signal intensity decreases. Arrays may be used past the manufacturer's expiration date.
- Perform spectral or spatial calibrations (see sections above).
- If the instrument will not be in use for more than 2 weeks, refer to the 3500xL User Guide for further information on proper care of the instrument.

PERFORMANCE CHECKS

A performance check will be done on an instrument if it has been moved and following repair or routine maintenance done by a qualified technician. The check must be completed prior to use for convicted offender sample analysis and consists of successfully running:

1. A ladder, a negative amplification control, and a positive amplification control (or other known sample) and
2. If the sensitivity of the instrument may have been affected (e.g. the optics were adjusted), a sensitivity check (see Quarterly maintenance).

For new instruments, a 1 base pair precision check is required in addition to the performance check detailed above.

GENEMAPPER® ID-X V1.4 SERVER MAINTENANCE

WEEKLY

1. Restart the computer
2. Review the statistics for the database (see the GeneMapper ID-X Administrator's Guide for the full procedure)

QUARTERLY

If ID-X is being used as an Expert System, run re-certification samples according to the NDIS Operational Procedures Manual

AS NEEDED

1. Review password policies
2. If using the audit system, review
3. If using the electronic signature system, review
4. Manage user accounts
5. Manage security groups
6. Manage profiles
7. Back up security settings after making changes
8. Back up and delete audit records when the record count is close to or exceeds 20,000.

TUTTNAUER TABLETOP AUTOCLAVE – INSTRUMENT MAINTENANCE

PRIOR TO USE

1. Fill the water reservoir
 - a. Remove the water reservoir cover.
 - b. Pour deionized water into the reservoir, through the opening on the top of the autoclave until it reaches the base of the safety valve holder (approximately 3 liters).

WEEKLY

1. Clean the door gasket
 - a. Clean the door gasket with a mild detergent, water and a soft cloth or sponge. The gasket should be clean and smooth.

MONTHLY

1. Change the water in the water reservoir
 - a. Ensure that the electrical cord is disconnected and there is no pressure in the autoclave.
 - b. To drain the reservoir, turn the drain valve counter clockwise to the open position. Place a bucket underneath to catch the flow of water. Fully drain the reservoir.
 - c. With a quart of tap water flush out the reservoir.
 - d. Turn the drain valve clockwise to the closed position and fill the reservoir with deionized water to just below the safety valve. Plug the electrical cord in.
2. Clean the water sensor
 - a. Wipe the sides and tip of the water sensor (located in the rear of the chamber) with a mild soapy solution.
 - b. It is very important the sides of the sensor are wiped clean.

QUARTERLY

1. Clean the air jet
 - a. Remove the water reservoir cover.
 - b. Clean the hole of the jet by manipulating the air trap wire back and forth ten (10) times.
 - c. It is preferred that the cleaning is done while the unit is running a cycle and under pressure, but it can also be done while the unit is idle.
2. Clean and descale the chamber, copper tubes, and reservoir using ChamberBrite™. Follow the procedure for cleaning the autoclave as outlined in the Tuttnauer Tabletop Autoclave maintenance manual.
3. Clean the trays and tray holder
 - a. Clean the tray holder and trays with detergent or non-abrasive stainless steel cleaner and water.
 - b. Rinse immediately to avoid staining the metal.

AS NEEDED (SEE PROCEDURES AS OUTLINED IN THE TUTTNAUER TABLETOP AUTOCLAVE MAINTENANCE MANUAL)

1. Clean and check the safety valve
2. Replace the door gasket
3. Clean the fan grid with compressed air from the inside outward
4. Clean the strainer
5. Inspect the locking device for wear

APPENDIX A: ADMINISTRATIVE PROCEDURES FOR PROCESSING OFFENDER SAMPLES

BLOOD SAMPLES

1. Check that the information on the data submission sheet is complete and matches that on the blood vial. Fill out "Received By" and "Date and Time Rec'd" sections.
2. See the [Processing of Convicted Offender Samples](#) section of this manual for more information about how blood samples are received and stored.

BUCCAL SAMPLES

1. Group sample envelopes by date received.
2. Group sample envelopes by submitting agency. If there is not a return address, group by postmark and open to determine agency. If there is no postmark, open envelope to determine agency.
3. Group sample envelopes by initials on the security seal.
 - a. If the seal is absent or tampered with, set the sample aside.
 - b. If the initials are absent, open envelope to verify that collector information is filled out. If not, set the sample aside.
4. Open envelopes one at a time keeping the groups together. Verify:
 - a. Offender information is filled out correctly, and a Washington State Identification (SID) Number is present and there is a qualifying offense listed on the submission card.
 - b. Collector information is complete:
 - i. Collector's signature is present.
 - If no signature is present, initial and date the "Collector's Signature" field after verifying that the collector has initialed the security seal on the envelope.
 - If neither signature nor initials are present, set the sample aside and then call to verify collector.
 - ii. The agency name is present and distinctive.
 - If the agency is not listed or the abbreviation is ambiguous and it can be determined from: the postmark, return address on the envelope or group of samples it was submitted with; fill in the appropriate information and initial and date the addition.
 - iii. If the collector information cannot be determined, set the sample aside and call for clarification.
 - c. The thumbprints are distinctive and the ridges can be clearly seen.
 - d. The DNA sample is obvious on the FTA[®] paper, i.e. pink paper has turned white.
 - e. For the multi-component kit, verify the convicted offender's name, SID number (an eight-digit number) and date of birth, if printed on the outer part of the FTA[®] card, corresponds to the information on the submission card.
 - f. If criteria in 4.a-4.e are met, skip to the [Assigning Lab Numbers](#) section below.
 - g. If 4.a is insufficient, look up the missing data on the convicted offender's criminal history, court records, or call the collecting agency for clarification. If an offense is determined to be non-qualifying, and the offender does not have a previous qualifying conviction, the sample must be rejected.

5. Agencies that submitted samples that were set aside because of a violation in step 3, 4.b.i, or 4.c-4.e should be contacted to request a recollection; emphasize that the entire kit needs to be redone. If a recollection is not possible, the sample may be accepted at the discretion of a CODIS staff member.
6. When a sample is rejected, fill out the rejection log with the pertinent information.

ASSIGNING LAB NUMBERS

1. Lab numbers depend upon lab location, the year of submission and the block of numbers assigned to the convicted offender program for that year by the CODIS Manager.
 - a. An example of a lab number assigned to a convicted offender is 103-015234.
 - i. The lab number begins with the lab code (Seattle = 1).
 - ii. The next two digits refer to the year (03 = 2003).
 - iii. The remaining six digits are the actual sample number; their start point is determined by the CODIS manager. Each successive sample is assigned the next number in the sequence.
2. For blood samples, based on the numbering rules above, write the next lab number, in sequential order, in the upper right corner of the submission sheet in the box labeled "Laboratory No."/"Crime Lab Use Only".
3. The remaining steps are for buccal samples. Begin with the presorted samples from the Receiving Samples section.
 - a. For the all-in-one sample card:
 - i. Number the card with the next available number in the sequence in the box labeled "for Crime Lab Use Only". Numbering may be done by hand or with pre-printed labels.
 - b. For the multi-component kit:
 - i. Number a 3½" x 6" clasp envelope with the next available number in the sequence at the top center of the envelope (the envelope flap should be to the left). Numbering may be done by hand or with pre-printed labels.
 - ii. Place the FTA card in the envelope.
 - iii. Number the corresponding submission card on the front, upper-right corner of the folded submission card.
4. Repeat with the remaining samples. Be sure to keep the numbers in order. It is acceptable to intermix both types of buccal collection kits.

ENTERING DATA INTO THE CONVICTED OFFENDER ENTRY FORM

1. For both blood and buccal samples, open Convicted Offender Entry Form (the user name and password are the same as for LIMS).
2. Make sure Caps Lock is on.
3. Select the proper entry type. When entering an all-in-one sample card, be sure to keep the FTA® paper covered.
 - a. Washington State Identification numbers (SIDs) are usually eight digits. When "SID Number" is selected, the entry field will contain a "WA" prefix.
 - b. If a SID number is unavailable, use the "Other Number" type and preface the number with one of the following abbreviations:

- i. Juvenile numbers are six digits or a letter and five digits and may be categorized as JUVIS, JRA or JV.
 - If JUVIS is specified, use JUV.
 - If JRA is specified, use JRA.
 - If nothing is specified, but juvenile can be determined by date of birth or facility, then use JV.
 - c. Social Security numbers should be abbreviated as SSN and entered without hyphens or spaces. Verify with the institution beforehand that this is the only ID available.
 - d. Department of Corrections numbers should be abbreviated as DOC. Verify with the institution beforehand that this is the only ID available.
 - e. Local Identification numbers should be abbreviated as Local ID. Verify with the institution beforehand that this is the only ID available.
 - f. Out-of-state Identification numbers should be prefaced with the state's two letter abbreviation. Verify with the institution beforehand that this is the only ID available.
 - g. FBI numbers shall be prefaced with FBI. Verify with the institution beforehand that this is the only ID available.
 - h. If no number is available indicate "No ID". Make all attempts to obtain a number from the submitting agency before this step.
4. Click Enter.
5. If using entry type 3.a, a dialogue box will appear containing the name, date of birth, and DNA status found in W2 for the entered SID number. If an error message appears, check the SID to verify that it was entered correctly and if it still does not work, set the sample aside to be looked up later.
 - a. Verify that name and date of birth match the submission card. Some offenders will have multiple names and/or dates of birth; the Convicted Offender Program will insert the first listed name or date of birth in W2 into the Convicted Offender Entry Form. For the sake of consistency, if there is more than one name and/or date of birth listed in the dialogue box, the first one in the list will be assigned to that SID number. If the submission card has an alias name and/or date of birth, note this on the card and write in what will be used (date and initial). If either the name or date of birth on the submission card is not listed in the dialogue box, set the sample aside to be looked up later.
 - b. Check DNA status.
 - i. If it says "DNA NOT Taken", Click OK.
 - ii. If it says "DNA Taken" stamp the following with a "Duplicate" stamp:
 - For blood samples, stamp the top of the submission sheet and outside of the zip top bag containing the blood sample.
 - For the all-in-one sample card, stamp the collector information side of the card.
 - For the multi-component kit, stamp the front center of the clasp envelope and the upper left corner of the open submission card.
 - iii. Click OK.
 - c. "Last Name", "First Name", "Middle Initial" (if available in W2), "Date of Birth", and "Individual Type" fields will populate automatically. Continue entry as follows:
 - i. Select Gender.
 - ii. Tab to "Offense Description". Enter offense from submission card. Common abbreviations are okay.
 - iii. Tab to "Collector's Institution". Enter institution from submission card.
 - iv. Tab to "Collector's Name". Select the representative that did the collection. If a particular representative is not on the list, they may be added by clicking on the plus (+) button.
 - Take off the Caps Lock.
 - Enter collector's last name, first name, and any other known collector information in the available fields.

- Type your initials and date added in the “Added by” field.
 - Click OK. (Make sure Caps Lock is back on).
 - v. Tab to “Request Date”. Enter date received.
 - vi. Tab to “Case Number” and enter the previously assigned lab number. Click OK.
 - vii. Verify appropriate barcode quantity is selected.
 - For blood samples, change this to 0.
 - For the all-in-one kit, change this to 1.
 - For the multi-component kit, leave as three.
 - viii. Click OK to complete the case entry and print barcode label(s).
6. If using entry type 3.b or 3.c, a new entry form will appear (“Individual Type” will be automatically set). Enter first and last name, middle initial, and date of birth. Continue as in steps 5.c.i – 5.c.x.

AFFIXING LABELS

1. Make sure the number on the label is the same as the number on the Submission Card; and for the multi-component kit, on the corresponding envelope.
2. Place a label on the “CRIME LAB USE ONLY” space on the Submission Card.
3. For the multi-component kit:
 - a. Verify that any information written on the FTA[®] card matches the corresponding Submission Card.
 - b. Place the next label on the outer cover of the corresponding FTA[®] card. Do not cover any convicted offender information that may be written on the flap; instead place the label on the back.
 - c. Place the last label over the number already present on the outside of the clasp envelope.
 - d. Place the FTA[®] and Submission Cards into the envelope.
4. File the sample in numerical order in the filing cabinet.

ADMINISTRATIVE ERRORS

Error	Description	Correction
Duplication of a case number	A particular case number is assigned to more than one sample.	LIMS will not allow duplicate entries of a case number. Upon discovery of such a duplicate, merely assign the next available case number to the submission card and sample envelope (initial if making a strikeout).
Mis-entering a case number	A case number is entered into the Convicted Offender form in the wrong format (e.g. 406-123456 or 1061-23456).	E-mail the LIMS Administrator with a request to delete the incorrect number. Wait for a response that the incorrect number was deleted from LIMS (otherwise a duplicate entry will be created). Ensure that the DNA Taken flag has been changed to no (unless the sample was already a duplicate). Re-enter the sample information with the correct number. Print and affix new labels.
Missing FTA card	An FTA card is not included with its corresponding submission card.	Set the card and envelope aside until the correct pairing can be established. If a match cannot be made, a request will be made of the agency to obtain another sample. Set the card/sample aside until the mismatch is ameliorated.
Switched FTA cards	An FTA card is incorrectly paired with a submission card.	Set the card and envelope aside until the correct pairing can be established. If a match cannot be made, a request may need to be made of the agency to obtain another sample. Set the card/sample aside until the mismatch is ameliorated.
Skipped case number	A case number is omitted in the sequence (e.g. 106-123456 → 106-123458), usually discovered by browsing stored submission cards.	Ensure that the missing number has not previously been assigned to a card that may have been removed from storage. Assign a new card & sample with the skipped case number. This should only be done for samples received in the same year.
Entering the wrong request date	The incorrect request/receipt date was entered into the Convicted Offender entry program.	Wait for the date to be updated in LIMS (usually a day after entry), then manually make the correction. Print and affix new labels.
Mislabeling	A submission card and/or sample envelope is labeled incorrectly.	If possible, remove the incorrect label(s) and determine to which card/envelope they belong. Affix to the appropriate card/envelope (reprint, if necessary, to maintain adhesive.) Print and affix the correct label(s) to the initial card/sample.

WORKSHEETS

CODIS CASE FILE FORMS

The worksheets approved for use in CODIS Casework can be found on the WSP Portal.

REFERENCES

- Applied Biosystems 3500/3500xL Genetic Analyzer User Guide, November 2012
- Applied Biosystems 7500/7500 Fast Real-Time PCR Systems Maintenance Guide, June 2010
- Applied Biosystems GeneAmp® PCR System 9700 User Guide and User Bulletin.
- Applied Biosystems GeneMapper® ID-X Software Version 1.0 Reference Guide, September 2007
- Applied Biosystems Quantifiler® Human and Y Human Male DNA Quantification Kits User Guide, January 2014
- Akane, A., Matsubara, K., Nakamura, H., Takahashi, S., and Kimura, K. Identification of the Heme Compound Copurified with DNA from Bloodstains, a Major Inhibitor of PCR Amplification, *J Forensic Sci* 1994;39:362-372.
- BSD600 Duet User Manual, Revision B, 12 August 2016
- Budowle, B., Chakraborty, R., Carmody, G., Mason, K. Source Attribution of a Forensic DNA Profile, *Forensic Science Communications* 2000; 2:1-6.
- Budowle, B., Moretti, T.R., Baumstark, A.L., Defenbaugh, D.A., and Keys, K.M. Population Data on the Thirteen CODIS Core Short Tandem Repeat Loci in African Americans, U.S. Caucasians, Hispanics, Bahamians, Jamaicans and Trinidadians, *J Forensic Sci* 1999;44:1277-1286.
- Budowle, B., Moretti, T.R., Keys, K.M., Koons, B.W., and Smerick, J.B. Validation Studies of the CTT STR Multiplex System, *J Forensic Sci* 1997;42(4):701-7.
- Budowle, B., Moretti, T.R., Buckleton, J.S. (2015), Erratum. *J Forensic Sci*. doi: 10.1111/1556-4029.12806 (pre-publication) and *J Forensic Sci* 2015;60(4)
- Budowle, B., Shea, B., Niezgoda, S., Chakraborty, R. CODIS STR Loci Data from 41 Sample Populations, *J Forensic Sci* 2001; 46:453-489.
- Budowle, B., Smerick, J.B., Keys, K.M., and Moretti, T.R. United States Population Data on the Multiplex Short Tandem Repeat Loci—HUMTHO1, TPOX and CSF1PO—and the Variable Number Tandem Repeat Locus D1S80, *J Forensic Sci* 1997;42(5):846-9. Published erratum appears in *J Forensic Sci* 1998;43(1):253.
- Chakraborty, R., Sample Size Requirements for Addressing Population Genetic Issues of Forensic Use of DNA Typing, *Human Biology* 1990;64:141-150.
- Comey, C.T., Koons, B.W., Presley, K.W., Smerick, J.B., Sobieralski, D.A., Stanley, D.M., and Baechtel, F.S. DNA Extraction Strategies for Amplified Fragment Length Polymorphism Analysis, *J Forensic Sci* 1994;39:1254-1269.
- Edwards, A., Civitello, A., Hammond, H.A., and Caskey, C.T. DNA Typing and Genetic Mapping with Trimeric and Tetrameric Tandem Repeats, *Am J Hum Genet* 1991;49:746-56.
- Ensenberger, M.G, Lenz, K.A, Matthies, L.K., Hadinoto, G.M., Schienman, J.E., Przech, A.J., Morganti, M.W., Renstrom, D.T., Baker, V.M., Gawrys, K.M., Hoogendoorn, M., Steffen, C.R., Martín, P., Alonso, A., Olso, H.R., Sprecher, C.J., and Storts, D.R. Developmental validation of the PowerPlex® Fusion 6C System, *Forensic Sci. Int. Genet.* 21 (2016) 134-144.

Fowler, J.C., Burgoyne, L.A., Scott, A.C., and Harding, H.W. Repetitive Deoxyribonucleic Acid (DNA) and Human Genome Variation—a Concise Review Relevant to Forensic Biology, J Forensic Sci 1988;33(5):1111-26.

Fregeau, C.J., Bowen, K.L., and Fourney, R.M. Validation of Highly Polymorphic Fluorescent Multiplex Short Tandem Repeat Systems Using Two Generations of DNA Sequencers, J Forensic Sci 1999;44(1):133-66.

Green, R.L., Roinestad, I.C., Boland, C. and Hennessy, L.K., Developmental Validation of the Quantifiler™ Real-Time PCR Kits for the Quantification of Human Nuclear DNA Samples, J. Forensic Sci. 50(4):809-25.

Guidelines for a Quality Assurance Program for DNA Analysis, Crime Lab Digest 1991;18:44-75.

Hammond, H.A., Li Jin, Zhong Y., Caskey, C.T., and Chakraborty R. Evaluation of 13 Short Tandem Repeat Loci for use in Personal Identification Applications, Am J Hum Genet 1994;55:175-89.

Hares, D. Selection and Implementation of Expanded CODIS Core Loci in the United States, Forensic Sci. Int. Genet. 17 (2015) 33-34.

Hochmeister, M.N., Budowle, B., Eisenberg, A., Borer, U.V., and Dirnhofer R. Using Multiplex PCR Amplification and Typing Kits for the Analysis of DNA Evidence in a Serial Killer Case, J Forensic Sci 1996;41(1):155-62.

LaFountain MJ, Schwartz MB, Svete PA, Walkinshaw MA, Buel E., TWGDAM validation of the AmpFISTR Profiler Plus and AmpFISTR COfiler STR multiplex systems using capillary electrophoresis. J Forensic Sci 2001;46:1191–8.

Lygo, J.E., Johnson, P.E., Holdaway, D.J., Woodroffe, S., Whitaker, J.P., Clayton, T.M., Kimpton, C.P., and Gill, P. The Validation of Short Tandem Repeat (STR) Loci for use in Forensic Casework, Int J Leg Med 1994;107:77-89.

Mansfield, E.S., Robertson, J.M., Vainer, M., Isenberg, A.R., Frazier, R.R., Ferguson, K., Chow, S., Harris, D.W., Barker, D.L., Gill, P.D., Budowle, B., and McCord, B.R. Analysis of Multiplexed Short Tandem Repeat (STR) Systems using Capillary Array Electrophoresis, Electrophoresis 1998;19(1):101-7.

National DNA Index System (NDIS): CODIS Standards for Acceptance of DNA Data at NDIS (1998).

National Research Council publications: DNA Technology in Forensic Science and The Evaluation of Forensic DNA Evidence. 1992 and 1996.

Promega DNA IQ™ System – Database Protocol (#TB297) and Small Sample Casework (#TB296) Technical Bulletins (9/16)

Promega PowerPlex® Fusion 6C Matrix Standard Technical Manual #TMD046 (10/15)

Promega PowerPlex® Fusion 6C System Technical Manual #TMD045 (6/16)

Promega PunchSolution™ Kit Technical Manual #TMD038 (9/16)

Quality Assurance Standards for DNA Databasing Laboratories, DAB/FBI (2011)

Reynolds, R., Sensabaugh, G., and Blake, E. Analysis of Genetic Markers in Forensic DNA Samples Using the Polymerase Chain Reaction, Analytical Chemistry 1991;63:2-15.

Saiki, R.K., Walsh, P.S., Levenson, C.H., and Erlich, H.A. Genetic Analysis of Amplified DNA with Immobilized Sequence-Specific Oligonucleotide Probes, Proc. Natl. Acad. Sci. USA 1989;86:6230-6234.

St.Claire III RL. Capillary Electrophoresis. Anal Chem 1996;68(12):569R-586R.

SWGDAM Interpretation Guidelines for Autosomal STR Typing by Forensic DNA Testing Laboratories, http://media.wix.com/ugd/4344b0_61b46a0e1a4c41ccb65f719a533b8e29.pdf

von Beroldingen, C.H., Blake, E., Higuchi, R., Sensabaugh, G., and Erlich, H.A. Application of PCR to the Analysis of Biological Evidence. PCR Technology Principles and Applications for DNA Amplification, Erlich, H.A., 1989; Stockton Press, New York.

Walsh, P.S., Erlich, H.A., and Higuchi, R. Preferential PCR Amplification of Alleles: Mechanisms and Solutions. PCR Methods and Applications 1992;1:241-250.

Walsh, P.S., Varlaro, J., and Reynolds, R. A Rapid Chemiluminescent Method for Quantitation of Human DNA, Nucleic Acids Research 1992;20:5061-5065.

Waye J.S., Presley, L.A., Budowle, B., Shutler, G.G., and Fourney, R.M. A Simple and Sensitive Method for Quantifying Human Genomic DNA in Forensic Specimen Extracts, Biotechniques 1989;7(8):852-5.